



Wednesday, December 19, 2018

Attn: Mr. Charles B. Sosik, P.G.
Environmental Business Consultants
1808 Middle Country Rd
Ridge NY 11961-2406

Project ID: 188 EAST 135TH ST BRONX NY
Sample ID#s: CC13976 - CC13983

This laboratory is in compliance with the NELAC requirements of procedures used except where indicated.

This report contains results for the parameters tested, under the sampling conditions described on the Chain Of Custody, as received by the laboratory. This report is incomplete unless all pages indicated in the pagination at the bottom of the page are included.

All soils, solids and sludges are reported on a dry weight basis unless otherwise noted in the sample comments.

A scanned version of the COC form accompanies the analytical report and is an exact duplicate of the original.

If you have any questions concerning this testing, please do not hesitate to contact Phoenix Client Services at ext. 200.

Sincerely yours,

A handwritten signature in black ink that reads "Phyllis Shiller". The signature is written in a cursive style.

Phyllis/Shiller
Laboratory Director

NELAC - #NY11301
CT Lab Registration #PH-0618
MA Lab Registration #M-CT007
ME Lab Registration #CT-007
NH Lab Registration #213693-A,B

NJ Lab Registration #CT-003
NY Lab Registration #11301
PA Lab Registration #68-03530
RI Lab Registration #63
UT Lab Registration #CT00007
VT Lab Registration #VT11301



Environmental Laboratories, Inc.
587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
Tel. (860) 645-1102 Fax (860) 645-0823



SDG Comments

December 19, 2018

SDG I.D.: GCC13976

8260 Volatile Organics:

1,2-Dibromoethane, 1,2,3 Trichloropropane, and 1,2-Dibromo-3-chloropropane do not meet NY TOGS GA criteria, these compounds are analyzed by GC/FID method 504 or 8011 to achieve this criteria.

Please be advised that the NY 375 soil criteria for chromium are based on hexavalent chromium and trivalent chromium.



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823

Analysis Report
 December 19, 2018

FOR: Attn: Mr. Charles B. Sosik, P.G.
 Environmental Business Consultants
 1808 Middle Country Rd
 Ridge NY 11961-2406

Sample Information

Matrix: SOIL
 Location Code: EBC
 Rush Request: 72 Hour
 P.O.#:

Custody Information

Collected by: DR
 Received by: SW
 Analyzed by: see "By" below

Date

12/12/18
 12/13/18

Time

8:15
 16:30

Laboratory Data

SDG ID: GCC13976
 Phoenix ID: CC13976

Project ID: 188 EAST 135TH ST BRONX NY
 Client ID: SB 1 (10-12')

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Silver	< 0.37	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Aluminum	5850	37		mg/Kg	10	12/15/18	CPP	SW6010C
Arsenic	15.8	0.75		mg/Kg	1	12/15/18	CPP	SW6010C
Barium	168	0.7		mg/Kg	1	12/15/18	EK	SW6010C
Beryllium	0.44	0.30		mg/Kg	1	12/15/18	CPP	SW6010C
Calcium	48200	37		mg/Kg	10	12/15/18	CPP	SW6010C
Cadmium	3.44	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Cobalt	10.3	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Chromium	46.3	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Copper	308	7.5		mg/kg	10	12/15/18	CPP	SW6010C
Iron	59600	37		mg/Kg	10	12/15/18	CPP	SW6010C
Mercury	1.29	0.15		mg/Kg	1	12/14/18	RS	SW7471B
Potassium	1120	7		mg/Kg	1	12/15/18	EK	SW6010C
Magnesium	14000	37		mg/Kg	10	12/15/18	CPP	SW6010C
Manganese	597	3.7		mg/Kg	10	12/15/18	CPP	SW6010C
Sodium	494	7		mg/Kg	1	12/15/18	CPP	SW6010C
Nickel	38.3	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Lead	662	7.5		mg/Kg	10	12/15/18	CPP	SW6010C
Antimony	15.1	3.7		mg/Kg	1	12/15/18	EK	SW6010C
Selenium	< 1.5	1.5		mg/Kg	1	12/15/18	CPP	SW6010C
Thallium	< 1.5	1.5		mg/Kg	1	12/15/18	CPP	SW6010C
Vanadium	56.3	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Zinc	720	7.5		mg/Kg	10	12/15/18	CPP	SW6010C
Percent Solid	87			%		12/13/18	AK	SW846-%Solid
Soil Extraction for SVOA	Completed					12/13/18	JJ/CK	SW3545A
Mercury Digestion	Completed					12/14/18	EV/EV	SW7471B
Total Metals Digest	Completed					12/14/18	M/AG	SW3050B

B

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,1-Trichloroethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,2,2-Tetrachloroethane	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
1,1,2-Trichloroethane	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloroethane	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloroethene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloropropene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,3-Trichlorobenzene	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
1,2,3-Trichloropropane	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
1,2,4-Trichlorobenzene	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
1,2,4-Trimethylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
1,2-Dibromo-3-chloropropane	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
1,2-Dibromoethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichlorobenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
1,2-Dichloroethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichloropropane	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
1,3,5-Trimethylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
1,3-Dichlorobenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
1,3-Dichloropropane	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
1,4-Dichlorobenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
2,2-Dichloropropane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
2-Chlorotoluene	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
2-Hexanone	ND	24	4.8	ug/Kg	1	12/15/18	JLI	SW8260C
2-Isopropyltoluene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
4-Chlorotoluene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
4-Methyl-2-pentanone	ND	24	4.8	ug/Kg	1	12/15/18	JLI	SW8260C
Acetone	9.8	JS 24	4.8	ug/Kg	1	12/15/18	JLI	SW8260C
Acrylonitrile	ND	9.7	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Benzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Bromobenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
Bromochloromethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Bromodichloromethane	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Bromoform	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Bromomethane	ND	4.8	1.9	ug/Kg	1	12/15/18	JLI	SW8260C
Carbon Disulfide	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Carbon tetrachloride	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Chlorobenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Chloroethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Chloroform	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Chloromethane	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
cis-1,2-Dichloroethene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
cis-1,3-Dichloropropene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Dibromochloromethane	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Dibromomethane	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Dichlorodifluoromethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Ethylbenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Hexachlorobutadiene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
Isopropylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
m&p-Xylene	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Methyl Ethyl Ketone	ND	29	4.8	ug/Kg	1	12/15/18	JLI	SW8260C
Methyl t-butyl ether (MTBE)	ND	9.7	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Methylene chloride	ND	4.8	4.8	ug/Kg	1	12/15/18	JLI	SW8260C
Naphthalene	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
n-Butylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
n-Propylbenzene	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
o-Xylene	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
p-Isopropyltoluene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
sec-Butylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
Styrene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
tert-Butylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
Tetrachloroethene	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Tetrahydrofuran (THF)	ND	9.7	2.4	ug/Kg	1	12/15/18	JLI	SW8260C
Toluene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,2-Dichloroethene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,3-Dichloropropene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,4-dichloro-2-butene	ND	640	160	ug/Kg	50	12/15/18	JLI	SW8260C
Trichloroethene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Trichlorofluoromethane	ND	4.8	0.97	ug/Kg	1	12/15/18	JLI	SW8260C
Trichlorotrifluoroethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Vinyl chloride	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
<u>QA/QC Surrogates</u>								
% 1,2-dichlorobenzene-d4	100			%	50	12/15/18	JLI	70 - 130 %
% Bromofluorobenzene	99			%	50	12/15/18	JLI	70 - 130 %
% Dibromofluoromethane	100			%	1	12/15/18	JLI	70 - 130 %
% Toluene-d8	94			%	1	12/15/18	JLI	70 - 130 %
<u>1,4-dioxane</u>								
1,4-dioxane	ND	72		ug/kg	1	12/15/18	JLI	SW8260C
<u>QA/QC Surrogates</u>								
% 1,2-dichlorobenzene-d4	121			%	1	12/15/18	JLI	70 - 130 %
% Bromofluorobenzene	73			%	1	12/15/18	JLI	70 - 130 %
% Toluene-d8	94			%	1	12/15/18	JLI	70 - 130 %
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	19		ug/Kg	1	12/15/18	JLI	SW8260C
Acrolein	ND	4.8		ug/Kg	1	12/15/18	JLI	SW8260C
Acrylonitrile	ND	19		ug/Kg	1	12/15/18	JLI	SW8260C
Tert-butyl alcohol	ND	97		ug/Kg	1	12/15/18	JLI	SW8260C
<u>Semivolatiles</u>								
1,2,4,5-Tetrachlorobenzene	ND	260	130	ug/Kg	1	12/14/18	WB	SW8270D
1,2,4-Trichlorobenzene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
1,2-Dichlorobenzene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
1,2-Diphenylhydrazine	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
1,3-Dichlorobenzene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
1,4-Dichlorobenzene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
2,4,5-Trichlorophenol	ND	260	210	ug/Kg	1	12/14/18	WB	SW8270D
2,4,6-Trichlorophenol	ND	190	120	ug/Kg	1	12/14/18	WB	SW8270D

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
2,4-Dichlorophenol	ND	190	130	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dimethylphenol	ND	260	94	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dinitrophenol	ND	260	260	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dinitrotoluene	ND	190	150	ug/Kg	1	12/14/18	WB	SW8270D
2,6-Dinitrotoluene	ND	190	120	ug/Kg	1	12/14/18	WB	SW8270D
2-Chloronaphthalene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Chlorophenol	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Methylnaphthalene	250	J 260	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Methylphenol (o-cresol)	ND	260	180	ug/Kg	1	12/14/18	WB	SW8270D
2-Nitroaniline	ND	260	260	ug/Kg	1	12/14/18	WB	SW8270D
2-Nitrophenol	ND	260	240	ug/Kg	1	12/14/18	WB	SW8270D
3&4-Methylphenol (m&p-cresol)	ND	260	150	ug/Kg	1	12/14/18	WB	SW8270D
3,3'-Dichlorobenzidine	ND	190	180	ug/Kg	1	12/14/18	WB	SW8270D
3-Nitroaniline	ND	380	760	ug/Kg	1	12/14/18	WB	SW8270D
4,6-Dinitro-2-methylphenol	ND	230	76	ug/Kg	1	12/14/18	WB	SW8270D
4-Bromophenyl phenyl ether	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
4-Chloro-3-methylphenol	ND	260	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Chloroaniline	ND	300	180	ug/Kg	1	12/14/18	WB	SW8270D
4-Chlorophenyl phenyl ether	ND	260	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Nitroaniline	ND	380	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Nitrophenol	ND	380	170	ug/Kg	1	12/14/18	WB	SW8270D
Acenaphthene	490	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Acenaphthylene	550	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Acetophenone	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Aniline	ND	300	300	ug/Kg	1	12/14/18	WB	SW8270D
Anthracene	1500	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Benz(a)anthracene	3400	260	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzidine	ND	380	220	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(a)pyrene	3200	190	120	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(b)fluoranthene	3000	260	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(ghi)perylene	1800	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(k)fluoranthene	2700	260	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzoic acid	ND	1900	760	ug/Kg	1	12/14/18	WB	SW8270D
Benzyl butyl phthalate	ND	260	97	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroethoxy)methane	ND	260	100	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroethyl)ether	ND	190	100	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroisopropyl)ether	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-ethylhexyl)phthalate	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Carbazole	360	190	150	ug/Kg	1	12/14/18	WB	SW8270D
Chrysene	3400	260	130	ug/Kg	1	12/14/18	WB	SW8270D
Dibenz(a,h)anthracene	720	190	120	ug/Kg	1	12/14/18	WB	SW8270D
Dibenzofuran	320	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Diethyl phthalate	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Dimethylphthalate	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Di-n-butylphthalate	ND	260	100	ug/Kg	1	12/14/18	WB	SW8270D
Di-n-octylphthalate	ND	260	97	ug/Kg	1	12/14/18	WB	SW8270D
Fluoranthene	7000	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Fluorene	470	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorobenzene	ND	190	110	ug/Kg	1	12/14/18	WB	SW8270D

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Hexachlorobutadiene	ND	260	140	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorocyclopentadiene	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Hexachloroethane	ND	190	110	ug/Kg	1	12/14/18	WB	SW8270D
Indeno(1,2,3-cd)pyrene	2100	260	130	ug/Kg	1	12/14/18	WB	SW8270D
Isophorone	ND	190	110	ug/Kg	1	12/14/18	WB	SW8270D
Naphthalene	210	J 260	110	ug/Kg	1	12/14/18	WB	SW8270D
Nitrobenzene	ND	190	130	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodimethylamine	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodi-n-propylamine	ND	190	120	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodiphenylamine	ND	260	150	ug/Kg	1	12/14/18	WB	SW8270D
Pentachloronitrobenzene	ND	260	140	ug/Kg	1	12/14/18	WB	SW8270D
Pentachlorophenol	ND	230	140	ug/Kg	1	12/14/18	WB	SW8270D
Phenanthrene	5000	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Phenol	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Pyrene	6500	260	130	ug/Kg	1	12/14/18	WB	SW8270D
Pyridine	ND	260	93	ug/Kg	1	12/14/18	WB	SW8270D
<u>QA/QC Surrogates</u>								
% 2,4,6-Tribromophenol	68			%	1	12/14/18	WB	30 - 130 %
% 2-Fluorobiphenyl	59			%	1	12/14/18	WB	30 - 130 %
% 2-Fluorophenol	50			%	1	12/14/18	WB	30 - 130 %
% Nitrobenzene-d5	58			%	1	12/14/18	WB	30 - 130 %
% Phenol-d5	58			%	1	12/14/18	WB	30 - 130 %
% Terphenyl-d14	60			%	1	12/14/18	WB	30 - 130 %
Field Extraction	Completed					12/12/18		SW5035A

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
-----------	--------	------------	-------------	-------	----------	-----------	----	-----------

1 = This parameter is not certified by the primary accrediting authority (NY NELAC) for this matrix. NY NELAC does not offer certification for all parameters at this time.

B = Present in blank, no bias suspected.

RL/PQL=Reporting/Practical Quantitation Level (Equivalent to NELAC LOQ, Limit of Quantitation) ND=Not Detected at RL/PQL
BRL=Below Reporting Level L=Biased Low J=Estimated Below RL LOD=Limit of Detection MDL=Method Detection Limit
QA/QC Surrogates: Surrogates are compounds (preceded with a %) added by the lab to determine analysis efficiency. Surrogate results(%) listed in the report are not "detected" compounds.

Comments:

Per 1.4.6 of EPA method 8270D, 1,2-Diphenylhydrazine is unstable and readily converts to Azobenzene. Azobenzene is used for the calibration of 1,2-Diphenylhydrazine.

Please be advised that the NY 375 soil criteria for chromium are based on hexavalent chromium and trivalent chromium.


Volatile Comment:

There was a suppression of the last internal standard in the low level analysis, all affected compounds are reported from the methanol preserved high level analysis which did not exhibit this interference.

All soils, solids and sludges are reported on a dry weight basis unless otherwise noted in the sample comments.

S - Laboratory solvent, contamination is possible.

If there are any questions regarding this data, please call Phoenix Client Services.
This report must not be reproduced except in full as defined by the attached chain of custody.



Phyllis Shiller, Laboratory Director

December 19, 2018

Reviewed and Released by: Greg Lawrence, Assistant Lab Director



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823

Analysis Report
 December 19, 2018

FOR: Attn: Mr. Charles B. Sosik, P.G.
 Environmental Business Consultants
 1808 Middle Country Rd
 Ridge NY 11961-2406

Sample Information

Matrix: SOIL
 Location Code: EBC
 Rush Request: 72 Hour
 P.O.#:

Custody Information

Collected by: DR
 Received by: SW
 Analyzed by: see "By" below

Date

12/12/18
 12/13/18

Time

11:45
 16:30

Laboratory Data

SDG ID: GCC13976
 Phoenix ID: CC13977

Project ID: 188 EAST 135TH ST BRONX NY
 Client ID: SB 2 (7-9')

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Silver	< 0.40	0.40		mg/Kg	1	12/15/18	CPP	SW6010C
Aluminum	9320	40		mg/Kg	10	12/15/18	CPP	SW6010C
Arsenic	8.39	0.80		mg/Kg	1	12/15/18	CPP	SW6010C
Barium	285	0.8		mg/Kg	1	12/15/18	EK	SW6010C
Beryllium	0.52	0.32		mg/Kg	1	12/15/18	CPP	SW6010C
Calcium	44500	40		mg/Kg	10	12/15/18	CPP	SW6010C
Cadmium	2.61	0.40		mg/Kg	1	12/15/18	CPP	SW6010C
Cobalt	9.08	0.40		mg/Kg	1	12/15/18	CPP	SW6010C
Chromium	23.2	0.40		mg/Kg	1	12/15/18	CPP	SW6010C
Copper	159	0.8		mg/kg	1	12/15/18	CPP	SW6010C
Iron	39300	40		mg/Kg	10	12/15/18	CPP	SW6010C
Mercury	13.0	1.4		mg/Kg	1	12/14/18	RS	SW7471B
Potassium	1840	8		mg/Kg	1	12/15/18	EK	SW6010C
Magnesium	4420	4.0		mg/Kg	1	12/15/18	CPP	SW6010C
Manganese	398	4.0		mg/Kg	10	12/15/18	CPP	SW6010C
Sodium	396	8		mg/Kg	1	12/15/18	CPP	SW6010C
Nickel	18.5	0.40		mg/Kg	1	12/15/18	CPP	SW6010C
Lead	959	8.0		mg/Kg	10	12/15/18	CPP	SW6010C
Antimony	< 4.0	4.0		mg/Kg	1	12/15/18	CPP	SW6010C
Selenium	< 1.6	1.6		mg/Kg	1	12/15/18	CPP	SW6010C
Thallium	< 1.6	1.6		mg/Kg	1	12/15/18	CPP	SW6010C
Vanadium	25.3	0.40		mg/Kg	1	12/15/18	CPP	SW6010C
Zinc	877	8.0		mg/Kg	10	12/15/18	CPP	SW6010C
Percent Solid	84			%		12/13/18	AK	SW846-%Solid
Soil Extraction for SVOA	Completed					12/13/18	JJ/CK	SW3545A
Mercury Digestion	Completed					12/14/18	EV/EV	SW7471B
Total Metals Digest	Completed					12/14/18	M/AG	SW3050B

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,1-Trichloroethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,2,2-Tetrachloroethane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,2-Trichloroethane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloroethane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloroethene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloropropene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,3-Trichlorobenzene	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,3-Trichloropropane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,4-Trichlorobenzene	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,4-Trimethylbenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dibromo-3-chloropropane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dibromoethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichlorobenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichloroethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichloropropane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
1,3,5-Trimethylbenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,3-Dichlorobenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
1,3-Dichloropropane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
1,4-Dichlorobenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
2,2-Dichloropropane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
2-Chlorotoluene	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
2-Hexanone	ND	24	4.8	ug/Kg	1	12/15/18	JLI	SW8260C
2-Isopropyltoluene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
4-Chlorotoluene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
4-Methyl-2-pentanone	ND	24	4.8	ug/Kg	1	12/15/18	JLI	SW8260C
Acetone	23	JS 24	4.8	ug/Kg	1	12/15/18	JLI	SW8260C
Acrylonitrile	ND	9.6	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Benzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Bromobenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Bromochloromethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Bromodichloromethane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Bromoform	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Bromomethane	ND	4.8	1.9	ug/Kg	1	12/15/18	JLI	SW8260C
Carbon Disulfide	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Carbon tetrachloride	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Chlorobenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Chloroethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Chloroform	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Chloromethane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
cis-1,2-Dichloroethene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
cis-1,3-Dichloropropene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Dibromochloromethane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Dibromomethane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Dichlorodifluoromethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Ethylbenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Hexachlorobutadiene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Isopropylbenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
m&p-Xylene	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Methyl Ethyl Ketone	ND	29	4.8	ug/Kg	1	12/15/18	JLI	SW8260C
Methyl t-butyl ether (MTBE)	ND	9.6	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Methylene chloride	ND	4.8	4.8	ug/Kg	1	12/15/18	JLI	SW8260C
Naphthalene	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
n-Butylbenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
n-Propylbenzene	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
o-Xylene	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
p-Isopropyltoluene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
sec-Butylbenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Styrene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
tert-Butylbenzene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Tetrachloroethene	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Tetrahydrofuran (THF)	ND	9.6	2.4	ug/Kg	1	12/15/18	JLI	SW8260C
Toluene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,2-Dichloroethene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,3-Dichloropropene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,4-dichloro-2-butene	ND	9.6	2.4	ug/Kg	1	12/15/18	JLI	SW8260C
Trichloroethene	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Trichlorofluoromethane	ND	4.8	0.96	ug/Kg	1	12/15/18	JLI	SW8260C
Trichlorotrifluoroethane	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
Vinyl chloride	ND	4.8	0.48	ug/Kg	1	12/15/18	JLI	SW8260C
<u>QA/QC Surrogates</u>								
% 1,2-dichlorobenzene-d4	100			%	1	12/15/18	JLI	70 - 130 %
% Bromofluorobenzene	95			%	1	12/15/18	JLI	70 - 130 %
% Dibromofluoromethane	95			%	1	12/15/18	JLI	70 - 130 %
% Toluene-d8	100			%	1	12/15/18	JLI	70 - 130 %
<u>1,4-dioxane</u>								
1,4-dioxane	ND	72		ug/kg	1	12/15/18	JLI	SW8260C
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	19		ug/Kg	1	12/15/18	JLI	SW8260C
Acrolein	ND	4.8		ug/Kg	1	12/15/18	JLI	SW8260C
Acrylonitrile	ND	19		ug/Kg	1	12/15/18	JLI	SW8260C
Tert-butyl alcohol	ND	96		ug/Kg	1	12/15/18	JLI	SW8260C
<u>Semivolatiles</u>								
1,2,4,5-Tetrachlorobenzene	ND	270	140	ug/Kg	1	12/14/18	WB	SW8270D
1,2,4-Trichlorobenzene	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
1,2-Dichlorobenzene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
1,2-Diphenylhydrazine	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
1,3-Dichlorobenzene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
1,4-Dichlorobenzene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
2,4,5-Trichlorophenol	ND	270	210	ug/Kg	1	12/14/18	WB	SW8270D
2,4,6-Trichlorophenol	ND	190	120	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dichlorophenol	ND	190	140	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dimethylphenol	ND	270	96	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dinitrophenol	ND	270	270	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dinitrotoluene	ND	190	150	ug/Kg	1	12/14/18	WB	SW8270D

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
2,6-Dinitrotoluene	ND	190	120	ug/Kg	1	12/14/18	WB	SW8270D
2-Chloronaphthalene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Chlorophenol	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Methylnaphthalene	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
2-Methylphenol (o-cresol)	ND	270	180	ug/Kg	1	12/14/18	WB	SW8270D
2-Nitroaniline	ND	270	270	ug/Kg	1	12/14/18	WB	SW8270D
2-Nitrophenol	ND	270	250	ug/Kg	1	12/14/18	WB	SW8270D
3&4-Methylphenol (m&p-cresol)	ND	270	150	ug/Kg	1	12/14/18	WB	SW8270D
3,3'-Dichlorobenzidine	ND	190	180	ug/Kg	1	12/14/18	WB	SW8270D
3-Nitroaniline	ND	390	770	ug/Kg	1	12/14/18	WB	SW8270D
4,6-Dinitro-2-methylphenol	ND	230	77	ug/Kg	1	12/14/18	WB	SW8270D
4-Bromophenyl phenyl ether	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
4-Chloro-3-methylphenol	ND	270	140	ug/Kg	1	12/14/18	WB	SW8270D
4-Chloroaniline	ND	310	180	ug/Kg	1	12/14/18	WB	SW8270D
4-Chlorophenyl phenyl ether	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Nitroaniline	ND	390	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Nitrophenol	ND	390	170	ug/Kg	1	12/14/18	WB	SW8270D
Acenaphthene	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Acenaphthylene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Acetophenone	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Aniline	ND	310	310	ug/Kg	1	12/14/18	WB	SW8270D
Anthracene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Benz(a)anthracene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzdine	ND	390	230	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(a)pyrene	ND	190	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(b)fluoranthene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(ghi)perylene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(k)fluoranthene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzoic acid	ND	1900	770	ug/Kg	1	12/14/18	WB	SW8270D
Benzyl butyl phthalate	ND	270	100	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroethoxy)methane	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroethyl)ether	ND	190	100	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroisopropyl)ether	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-ethylhexyl)phthalate	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Carbazole	ND	190	150	ug/Kg	1	12/14/18	WB	SW8270D
Chrysene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Dibenz(a,h)anthracene	ND	190	130	ug/Kg	1	12/14/18	WB	SW8270D
Dibenzofuran	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Diethyl phthalate	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Dimethylphthalate	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Di-n-butylphthalate	ND	270	100	ug/Kg	1	12/14/18	WB	SW8270D
Di-n-octylphthalate	ND	270	100	ug/Kg	1	12/14/18	WB	SW8270D
Fluoranthene	320	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Fluorene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorobenzene	ND	190	110	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorobutadiene	ND	270	140	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorocyclopentadiene	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Hexachloroethane	ND	190	120	ug/Kg	1	12/14/18	WB	SW8270D
Indeno(1,2,3-cd)pyrene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D

1

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Isophorone	ND	190	110	ug/Kg	1	12/14/18	WB	SW8270D
Naphthalene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Nitrobenzene	ND	190	140	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodimethylamine	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodi-n-propylamine	ND	190	130	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodiphenylamine	ND	270	150	ug/Kg	1	12/14/18	WB	SW8270D
Pentachloronitrobenzene	ND	270	140	ug/Kg	1	12/14/18	WB	SW8270D
Pentachlorophenol	ND	230	150	ug/Kg	1	12/14/18	WB	SW8270D
Phenanthrene	300	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Phenol	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Pyrene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Pyridine	ND	270	95	ug/Kg	1	12/14/18	WB	SW8270D
QA/QC Surrogates								
% 2,4,6-Tribromophenol	73			%	1	12/14/18	WB	30 - 130 %
% 2-Fluorobiphenyl	63			%	1	12/14/18	WB	30 - 130 %
% 2-Fluorophenol	50			%	1	12/14/18	WB	30 - 130 %
% Nitrobenzene-d5	60			%	1	12/14/18	WB	30 - 130 %
% Phenol-d5	61			%	1	12/14/18	WB	30 - 130 %
% Terphenyl-d14	65			%	1	12/14/18	WB	30 - 130 %
Field Extraction	Completed					12/12/18		SW5035A

1 = This parameter is not certified by the primary accrediting authority (NY NELAC) for this matrix. NY NELAC does not offer certification for all parameters at this time.

B = Present in blank, no bias suspected.

RL/PQL=Reporting/Practical Quantitation Level (Equivalent to NELAC LOQ, Limit of Quantitation) ND=Not Detected at RL/PQL
 BRL=Below Reporting Level L=Biased Low J=Estimated Below RL LOD=Limit of Detection MDL=Method Detection Limit

QA/QC Surrogates: Surrogates are compounds (preceeded with a %) added by the lab to determine analysis efficiency. Surrogate results(%) listed in the report are not "detected" compounds.

Comments:

Per 1.4.6 of EPA method 8270D, 1,2-Diphenylhydrazine is unstable and readily converts to Azobenzene. Azobenzene is used for the calibration of 1,2-Diphenylhydrazine.


Please be advised that the NY 375 soil criteria for chromium are based on hexavalent chromium and trivalent chromium.

All soils, solids and sludges are reported on a dry weight basis unless otherwise noted in the sample comments.

S - Laboratory solvent, contamination is possible.

If there are any questions regarding this data, please call Phoenix Client Services.

This report must not be reproduced except in full as defined by the attached chain of custody.



Phyllis Shiller, Laboratory Director

December 19, 2018

Reviewed and Released by: Greg Lawrence, Assistant Lab Director



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823

Analysis Report
 December 19, 2018

FOR: Attn: Mr. Charles B. Sosik, P.G.
 Environmental Business Consultants
 1808 Middle Country Rd
 Ridge NY 11961-2406

Sample Information

Matrix: SOIL
 Location Code: EBC
 Rush Request: 72 Hour
 P.O.#:

Custody Information

Collected by: DR
 Received by: SW
 Analyzed by: see "By" below

Date

12/12/18
 12/13/18

Time

10:30
 16:30

Laboratory Data

SDG ID: GCC13976
 Phoenix ID: CC13978

Project ID: 188 EAST 135TH ST BRONX NY
 Client ID: SB 3 (7-9')

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Silver	< 0.37	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Aluminum	9190	37		mg/Kg	10	12/15/18	CPP	SW6010C
Arsenic	8.23	0.74		mg/Kg	1	12/15/18	CPP	SW6010C
Barium	203	0.7		mg/Kg	1	12/15/18	EK	SW6010C
Beryllium	0.56	0.29		mg/Kg	1	12/15/18	CPP	SW6010C
Calcium	6280	3.7		mg/Kg	1	12/15/18	CPP	SW6010C
Cadmium	0.90	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Cobalt	8.51	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Chromium	34.8	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Copper	63.9	0.7		mg/kg	1	12/15/18	CPP	SW6010C
Iron	25000	37		mg/Kg	10	12/15/18	CPP	SW6010C
Mercury	0.53	0.14		mg/Kg	1	12/14/18	RS	SW7471B
Potassium	1310	7		mg/Kg	1	12/15/18	EK	SW6010C
Magnesium	3000	3.7		mg/Kg	1	12/15/18	CPP	SW6010C
Manganese	228	3.7		mg/Kg	10	12/15/18	CPP	SW6010C
Sodium	375	7		mg/Kg	1	12/15/18	CPP	SW6010C
Nickel	43.2	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Lead	382	7.4		mg/Kg	10	12/15/18	CPP	SW6010C
Antimony	< 3.7	3.7		mg/Kg	1	12/15/18	CPP	SW6010C
Selenium	< 1.5	1.5		mg/Kg	1	12/15/18	CPP	SW6010C
Thallium	< 1.5	1.5		mg/Kg	1	12/15/18	CPP	SW6010C
Vanadium	23.5	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Zinc	195	7.4		mg/Kg	10	12/15/18	CPP	SW6010C
Percent Solid	87			%		12/13/18	AK	SW846-%Solid
Soil Extraction for SVOA	Completed					12/13/18	JJ/CK	SW3545A
Mercury Digestion	Completed					12/14/18	EV/EV	SW7471B
Total Metals Digest	Completed					12/14/18	M/AG	SW3050B

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,1-Trichloroethane	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,2,2-Tetrachloroethane	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
1,1,2-Trichloroethane	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloroethane	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloroethene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloropropene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,3-Trichlorobenzene	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
1,2,3-Trichloropropane	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
1,2,4-Trichlorobenzene	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
1,2,4-Trimethylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
1,2-Dibromo-3-chloropropane	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
1,2-Dibromoethane	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichlorobenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
1,2-Dichloroethane	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichloropropane	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
1,3,5-Trimethylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
1,3-Dichlorobenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
1,3-Dichloropropane	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
1,4-Dichlorobenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
2,2-Dichloropropane	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
2-Chlorotoluene	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
2-Hexanone	ND	31	6.3	ug/Kg	1	12/15/18	JLI	SW8260C
2-Isopropyltoluene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
4-Chlorotoluene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
4-Methyl-2-pentanone	ND	31	6.3	ug/Kg	1	12/15/18	JLI	SW8260C
Acetone	23	JS 31	6.3	ug/Kg	1	12/15/18	JLI	SW8260C
Acrylonitrile	ND	13	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Benzene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
Bromobenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
Bromochloromethane	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
Bromodichloromethane	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Bromoform	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Bromomethane	ND	6.3	2.5	ug/Kg	1	12/15/18	JLI	SW8260C
Carbon Disulfide	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Carbon tetrachloride	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Chlorobenzene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
Chloroethane	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
Chloroform	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
Chloromethane	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
cis-1,2-Dichloroethene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
cis-1,3-Dichloropropene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
Dibromochloromethane	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Dibromomethane	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Dichlorodifluoromethane	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
Ethylbenzene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
Hexachlorobutadiene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
Isopropylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
m&p-Xylene	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Methyl Ethyl Ketone	ND	38	6.3	ug/Kg	1	12/15/18	JLI	SW8260C
Methyl t-butyl ether (MTBE)	ND	13	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Methylene chloride	ND	6.3	6.3	ug/Kg	1	12/15/18	JLI	SW8260C
Naphthalene	71	J 320	64	ug/Kg	50	12/15/18	JLI	SW8260C
n-Butylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
n-Propylbenzene	ND	320	64	ug/Kg	50	12/15/18	JLI	SW8260C
o-Xylene	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
p-Isopropyltoluene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
sec-Butylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
Styrene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
tert-Butylbenzene	ND	320	32	ug/Kg	50	12/15/18	JLI	SW8260C
Tetrachloroethene	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Tetrahydrofuran (THF)	ND	13	3.1	ug/Kg	1	12/15/18	JLI	SW8260C
Toluene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,2-Dichloroethene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,3-Dichloropropene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,4-dichloro-2-butene	ND	640	160	ug/Kg	50	12/15/18	JLI	SW8260C
Trichloroethene	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
Trichlorofluoromethane	ND	6.3	1.3	ug/Kg	1	12/15/18	JLI	SW8260C
Trichlorotrifluoroethane	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
Vinyl chloride	ND	6.3	0.63	ug/Kg	1	12/15/18	JLI	SW8260C
<u>QA/QC Surrogates</u>								
% 1,2-dichlorobenzene-d4	100			%	50	12/15/18	JLI	70 - 130 %
% Bromofluorobenzene	99			%	50	12/15/18	JLI	70 - 130 %
% Dibromofluoromethane	98			%	1	12/15/18	JLI	70 - 130 %
% Toluene-d8	95			%	1	12/15/18	JLI	70 - 130 %
<u>1,4-dioxane</u>								
1,4-dioxane	ND	94		ug/kg	1	12/15/18	JLI	SW8260C
<u>QA/QC Surrogates</u>								
% 1,2-dichlorobenzene-d4	105			%	1	12/15/18	JLI	70 - 130 %
% Bromofluorobenzene	82			%	1	12/15/18	JLI	70 - 130 %
% Toluene-d8	95			%	1	12/15/18	JLI	70 - 130 %
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	25		ug/Kg	1	12/15/18	JLI	SW8260C
Acrolein	ND	6.3		ug/Kg	1	12/15/18	JLI	SW8260C
Acrylonitrile	ND	25		ug/Kg	1	12/15/18	JLI	SW8260C
Tert-butyl alcohol	ND	130		ug/Kg	1	12/15/18	JLI	SW8260C
<u>Semivolatiles</u>								
1,2,4,5-Tetrachlorobenzene	ND	260	130	ug/Kg	1	12/14/18	WB	SW8270D
1,2,4-Trichlorobenzene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
1,2-Dichlorobenzene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
1,2-Diphenylhydrazine	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
1,3-Dichlorobenzene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
1,4-Dichlorobenzene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
2,4,5-Trichlorophenol	ND	260	210	ug/Kg	1	12/14/18	WB	SW8270D
2,4,6-Trichlorophenol	ND	190	120	ug/Kg	1	12/14/18	WB	SW8270D

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
2,4-Dichlorophenol	ND	190	130	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dimethylphenol	ND	260	93	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dinitrophenol	ND	260	260	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dinitrotoluene	ND	190	150	ug/Kg	1	12/14/18	WB	SW8270D
2,6-Dinitrotoluene	ND	190	120	ug/Kg	1	12/14/18	WB	SW8270D
2-Chloronaphthalene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Chlorophenol	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Methylnaphthalene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Methylphenol (o-cresol)	ND	260	180	ug/Kg	1	12/14/18	WB	SW8270D
2-Nitroaniline	ND	260	260	ug/Kg	1	12/14/18	WB	SW8270D
2-Nitrophenol	ND	260	240	ug/Kg	1	12/14/18	WB	SW8270D
3&4-Methylphenol (m&p-cresol)	ND	260	150	ug/Kg	1	12/14/18	WB	SW8270D
3,3'-Dichlorobenzidine	ND	190	180	ug/Kg	1	12/14/18	WB	SW8270D
3-Nitroaniline	ND	380	750	ug/Kg	1	12/14/18	WB	SW8270D
4,6-Dinitro-2-methylphenol	ND	230	75	ug/Kg	1	12/14/18	WB	SW8270D
4-Bromophenyl phenyl ether	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
4-Chloro-3-methylphenol	ND	260	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Chloroaniline	ND	300	180	ug/Kg	1	12/14/18	WB	SW8270D
4-Chlorophenyl phenyl ether	ND	260	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Nitroaniline	ND	380	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Nitrophenol	ND	380	170	ug/Kg	1	12/14/18	WB	SW8270D
Acenaphthene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Acenaphthylene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Acetophenone	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Aniline	ND	300	300	ug/Kg	1	12/14/18	WB	SW8270D
Anthracene	140	J 260	120	ug/Kg	1	12/14/18	WB	SW8270D
Benz(a)anthracene	430	260	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzidine	ND	380	220	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(a)pyrene	470	190	120	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(b)fluoranthene	400	260	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(ghi)perylene	330	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(k)fluoranthene	380	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Benzoic acid	ND	1900	750	ug/Kg	1	12/14/18	WB	SW8270D
Benzyl butyl phthalate	ND	260	97	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroethoxy)methane	ND	260	100	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroethyl)ether	ND	190	100	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroisopropyl)ether	ND	260	100	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-ethylhexyl)phthalate	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Carbazole	ND	190	150	ug/Kg	1	12/14/18	WB	SW8270D
Chrysene	450	260	130	ug/Kg	1	12/14/18	WB	SW8270D
Dibenz(a,h)anthracene	ND	190	120	ug/Kg	1	12/14/18	WB	SW8270D
Dibenzofuran	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Diethyl phthalate	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Dimethylphthalate	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Di-n-butylphthalate	ND	260	100	ug/Kg	1	12/14/18	WB	SW8270D
Di-n-octylphthalate	ND	260	97	ug/Kg	1	12/14/18	WB	SW8270D
Fluoranthene	780	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Fluorene	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorobenzene	ND	190	110	ug/Kg	1	12/14/18	WB	SW8270D

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Hexachlorobutadiene	ND	260	140	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorocyclopentadiene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Hexachloroethane	ND	190	110	ug/Kg	1	12/14/18	WB	SW8270D
Indeno(1,2,3-cd)pyrene	330	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Isophorone	ND	190	110	ug/Kg	1	12/14/18	WB	SW8270D
Naphthalene	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Nitrobenzene	ND	190	130	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodimethylamine	ND	260	110	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodi-n-propylamine	ND	190	120	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodiphenylamine	ND	260	140	ug/Kg	1	12/14/18	WB	SW8270D
Pentachloronitrobenzene	ND	260	140	ug/Kg	1	12/14/18	WB	SW8270D
Pentachlorophenol	ND	230	140	ug/Kg	1	12/14/18	WB	SW8270D
Phenanthrene	520	260	110	ug/Kg	1	12/14/18	WB	SW8270D
Phenol	ND	260	120	ug/Kg	1	12/14/18	WB	SW8270D
Pyrene	740	260	130	ug/Kg	1	12/14/18	WB	SW8270D
Pyridine	ND	260	92	ug/Kg	1	12/14/18	WB	SW8270D
<u>QA/QC Surrogates</u>								
% 2,4,6-Tribromophenol	79			%	1	12/14/18	WB	30 - 130 %
% 2-Fluorobiphenyl	61			%	1	12/14/18	WB	30 - 130 %
% 2-Fluorophenol	55			%	1	12/14/18	WB	30 - 130 %
% Nitrobenzene-d5	57			%	1	12/14/18	WB	30 - 130 %
% Phenol-d5	61			%	1	12/14/18	WB	30 - 130 %
% Terphenyl-d14	60			%	1	12/14/18	WB	30 - 130 %
Field Extraction	Completed					12/12/18		SW5035A

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
-----------	--------	------------	-------------	-------	----------	-----------	----	-----------

1 = This parameter is not certified by the primary accrediting authority (NY NELAC) for this matrix. NY NELAC does not offer certification for all parameters at this time.

B = Present in blank, no bias suspected.

RL/PQL=Reporting/Practical Quantitation Level (Equivalent to NELAC LOQ, Limit of Quantitation) ND=Not Detected at RL/PQL
BRL=Below Reporting Level L=Biased Low J=Estimated Below RL LOD=Limit of Detection MDL=Method Detection Limit
QA/QC Surrogates: Surrogates are compounds (preceded with a %) added by the lab to determine analysis efficiency. Surrogate results(%) listed in the report are not "detected" compounds.

Comments:

Per 1.4.6 of EPA method 8270D, 1,2-Diphenylhydrazine is unstable and readily converts to Azobenzene. Azobenzene is used for the calibration of 1,2-Diphenylhydrazine.

Please be advised that the NY 375 soil criteria for chromium are based on hexavalent chromium and trivalent chromium.

Volatile Comment:

There was a suppression of the last internal standard in the low level analysis, all affected compounds are reported from the methanol preserved high level analysis which did not exhibit this interference.

All soils, solids and sludges are reported on a dry weight basis unless otherwise noted in the sample comments.

S - Laboratory solvent, contamination is possible.

If there are any questions regarding this data, please call Phoenix Client Services.
This report must not be reproduced except in full as defined by the attached chain of custody.

Phyllis Shiller, Laboratory Director

December 19, 2018

Reviewed and Released by: Greg Lawrence, Assistant Lab Director



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823

Analysis Report
 December 19, 2018

FOR: Attn: Mr. Charles B. Sosik, P.G.
 Environmental Business Consultants
 1808 Middle Country Rd
 Ridge NY 11961-2406

Sample Information

Matrix: SOIL
 Location Code: EBC
 Rush Request: 72 Hour
 P.O.#:

Custody Information

Collected by: DR
 Received by: SW
 Analyzed by: see "By" below

Date

12/12/18
 12/13/18

Time

9:45
 16:30

Laboratory Data

SDG ID: GCC13976
 Phoenix ID: CC13979

Project ID: 188 EAST 135TH ST BRONX NY
 Client ID: SB 4 (8-10')

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Silver	< 0.38	0.38		mg/Kg	1	12/15/18	CPP	SW6010C
Aluminum	9910	38		mg/Kg	10	12/15/18	CPP	SW6010C
Arsenic	5.21	0.76		mg/Kg	1	12/15/18	CPP	SW6010C
Barium	71.1	0.8		mg/Kg	1	12/15/18	EK	SW6010C
Beryllium	0.49	0.30		mg/Kg	1	12/15/18	CPP	SW6010C
Calcium	1540	3.8		mg/Kg	1	12/15/18	CPP	SW6010C
Cadmium	0.43	0.38		mg/Kg	1	12/15/18	CPP	SW6010C
Cobalt	10.7	0.38		mg/Kg	1	12/15/18	CPP	SW6010C
Chromium	17.4	0.38		mg/Kg	1	12/15/18	CPP	SW6010C
Copper	26.8	0.8		mg/kg	1	12/15/18	CPP	SW6010C
Iron	20000	38		mg/Kg	10	12/15/18	CPP	SW6010C
Mercury	0.24	0.13		mg/Kg	1	12/14/18	RS	SW7471B
Potassium	2030	8		mg/Kg	1	12/15/18	EK	SW6010C
Magnesium	2980	3.8		mg/Kg	1	12/15/18	CPP	SW6010C
Manganese	293	3.8		mg/Kg	10	12/15/18	CPP	SW6010C
Sodium	125	8		mg/Kg	1	12/15/18	CPP	SW6010C
Nickel	16.3	0.38		mg/Kg	1	12/15/18	CPP	SW6010C
Lead	62.6	0.8		mg/Kg	1	12/15/18	CPP	SW6010C
Antimony	< 3.8	3.8		mg/Kg	1	12/15/18	CPP	SW6010C
Selenium	< 1.5	1.5		mg/Kg	1	12/15/18	CPP	SW6010C
Thallium	< 1.5	1.5		mg/Kg	1	12/15/18	CPP	SW6010C
Vanadium	26.3	0.38		mg/Kg	1	12/15/18	CPP	SW6010C
Zinc	85.6	0.8		mg/Kg	1	12/15/18	CPP	SW6010C
Percent Solid	90			%		12/13/18	AK	SW846-%Solid
Soil Extraction for SVOA	Completed					12/13/18	JJ/CK	SW3545A
Mercury Digestion	Completed					12/14/18	EV/EV	SW7471B
Total Metals Digest	Completed					12/14/18	M/AG	SW3050B

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,1-Trichloroethane	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,2,2-Tetrachloroethane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,2-Trichloroethane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloroethane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloroethene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloropropene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,3-Trichlorobenzene	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,3-Trichloropropane	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,4-Trichlorobenzene	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,4-Trimethylbenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dibromo-3-chloropropane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dibromoethane	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichlorobenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichloroethane	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichloropropane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
1,3,5-Trimethylbenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
1,3-Dichlorobenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
1,3-Dichloropropane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
1,4-Dichlorobenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
2,2-Dichloropropane	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
2-Chlorotoluene	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
2-Hexanone	ND	28	5.7	ug/Kg	1	12/15/18	JLI	SW8260C
2-Isopropyltoluene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
4-Chlorotoluene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
4-Methyl-2-pentanone	ND	28	5.7	ug/Kg	1	12/15/18	JLI	SW8260C
Acetone	10	JS 28	5.7	ug/Kg	1	12/15/18	JLI	SW8260C
Acrylonitrile	ND	11	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Benzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Bromobenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Bromochloromethane	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Bromodichloromethane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Bromoform	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Bromomethane	ND	5.7	2.3	ug/Kg	1	12/15/18	JLI	SW8260C
Carbon Disulfide	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Carbon tetrachloride	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Chlorobenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Chloroethane	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Chloroform	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Chloromethane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
cis-1,2-Dichloroethene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
cis-1,3-Dichloropropene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Dibromochloromethane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Dibromomethane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Dichlorodifluoromethane	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Ethylbenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Hexachlorobutadiene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Isopropylbenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
m&p-Xylene	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Methyl Ethyl Ketone	ND	34	5.7	ug/Kg	1	12/15/18	JLI	SW8260C
Methyl t-butyl ether (MTBE)	ND	11	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Methylene chloride	ND	5.7	5.7	ug/Kg	1	12/15/18	JLI	SW8260C
Naphthalene	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
n-Butylbenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
n-Propylbenzene	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
o-Xylene	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
p-Isopropyltoluene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
sec-Butylbenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Styrene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
tert-Butylbenzene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Tetrachloroethene	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Tetrahydrofuran (THF)	ND	11	2.8	ug/Kg	1	12/15/18	JLI	SW8260C
Toluene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,2-Dichloroethene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,3-Dichloropropene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,4-dichloro-2-butene	ND	11	2.8	ug/Kg	1	12/15/18	JLI	SW8260C
Trichloroethene	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Trichlorofluoromethane	ND	5.7	1.1	ug/Kg	1	12/15/18	JLI	SW8260C
Trichlorotrifluoroethane	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
Vinyl chloride	ND	5.7	0.57	ug/Kg	1	12/15/18	JLI	SW8260C
<u>QA/QC Surrogates</u>								
% 1,2-dichlorobenzene-d4	112			%	1	12/15/18	JLI	70 - 130 %
% Bromofluorobenzene	90			%	1	12/15/18	JLI	70 - 130 %
% Dibromofluoromethane	94			%	1	12/15/18	JLI	70 - 130 %
% Toluene-d8	98			%	1	12/15/18	JLI	70 - 130 %
<u>1,4-dioxane</u>								
1,4-dioxane	ND	85		ug/kg	1	12/15/18	JLI	SW8260C
<u>QA/QC Surrogates</u>								
% 1,2-dichlorobenzene-d4	112			%	1	12/15/18	JLI	70 - 130 %
% Bromofluorobenzene	90			%	1	12/15/18	JLI	70 - 130 %
% Toluene-d8	98			%	1	12/15/18	JLI	70 - 130 %
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	23		ug/Kg	1	12/15/18	JLI	SW8260C
Acrolein	ND	5.7		ug/Kg	1	12/15/18	JLI	SW8260C
Acrylonitrile	ND	23		ug/Kg	1	12/15/18	JLI	SW8260C
Tert-butyl alcohol	ND	110		ug/Kg	1	12/15/18	JLI	SW8260C
<u>Semivolatiles</u>								
1,2,4,5-Tetrachlorobenzene	ND	250	130	ug/Kg	1	12/14/18	WB	SW8270D
1,2,4-Trichlorobenzene	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
1,2-Dichlorobenzene	ND	250	100	ug/Kg	1	12/14/18	WB	SW8270D
1,2-Diphenylhydrazine	ND	250	120	ug/Kg	1	12/14/18	WB	SW8270D
1,3-Dichlorobenzene	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
1,4-Dichlorobenzene	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
2,4,5-Trichlorophenol	ND	250	200	ug/Kg	1	12/14/18	WB	SW8270D
2,4,6-Trichlorophenol	ND	180	120	ug/Kg	1	12/14/18	WB	SW8270D

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
2,4-Dichlorophenol	ND	180	130	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dimethylphenol	ND	250	89	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dinitrophenol	ND	250	250	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dinitrotoluene	ND	180	140	ug/Kg	1	12/14/18	WB	SW8270D
2,6-Dinitrotoluene	ND	180	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Chloronaphthalene	ND	250	100	ug/Kg	1	12/14/18	WB	SW8270D
2-Chlorophenol	ND	250	100	ug/Kg	1	12/14/18	WB	SW8270D
2-Methylnaphthalene	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Methylphenol (o-cresol)	ND	250	170	ug/Kg	1	12/14/18	WB	SW8270D
2-Nitroaniline	ND	250	250	ug/Kg	1	12/14/18	WB	SW8270D
2-Nitrophenol	ND	250	230	ug/Kg	1	12/14/18	WB	SW8270D
3&4-Methylphenol (m&p-cresol)	ND	250	140	ug/Kg	1	12/14/18	WB	SW8270D
3,3'-Dichlorobenzidine	ND	180	170	ug/Kg	1	12/14/18	WB	SW8270D
3-Nitroaniline	ND	360	720	ug/Kg	1	12/14/18	WB	SW8270D
4,6-Dinitro-2-methylphenol	ND	220	72	ug/Kg	1	12/14/18	WB	SW8270D
4-Bromophenyl phenyl ether	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
4-Chloro-3-methylphenol	ND	250	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Chloroaniline	ND	290	170	ug/Kg	1	12/14/18	WB	SW8270D
4-Chlorophenyl phenyl ether	ND	250	120	ug/Kg	1	12/14/18	WB	SW8270D
4-Nitroaniline	ND	360	120	ug/Kg	1	12/14/18	WB	SW8270D
4-Nitrophenol	ND	360	160	ug/Kg	1	12/14/18	WB	SW8270D
Acenaphthene	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
Acenaphthylene	ND	250	100	ug/Kg	1	12/14/18	WB	SW8270D
Acetophenone	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
Aniline	ND	290	290	ug/Kg	1	12/14/18	WB	SW8270D
Anthracene	380	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Benz(a)anthracene	700	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Benzidine	ND	360	210	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(a)pyrene	640	180	120	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(b)fluoranthene	530	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(ghi)perylene	330	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(k)fluoranthene	500	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Benzoic acid	ND	1800	720	ug/Kg	1	12/14/18	WB	SW8270D
Benzyl butyl phthalate	ND	250	93	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroethoxy)methane	ND	250	99	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroethyl)ether	ND	180	97	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroisopropyl)ether	ND	250	100	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-ethylhexyl)phthalate	ND	250	100	ug/Kg	1	12/14/18	WB	SW8270D
Carbazole	ND	180	140	ug/Kg	1	12/14/18	WB	SW8270D
Chrysene	680	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Dibenz(a,h)anthracene	ND	180	120	ug/Kg	1	12/14/18	WB	SW8270D
Dibenzofuran	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
Diethyl phthalate	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
Dimethylphthalate	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
Di-n-butylphthalate	ND	250	96	ug/Kg	1	12/14/18	WB	SW8270D
Di-n-octylphthalate	ND	250	93	ug/Kg	1	12/14/18	WB	SW8270D
Fluoranthene	1700	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Fluorene	ND	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorobenzene	ND	180	110	ug/Kg	1	12/14/18	WB	SW8270D

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Hexachlorobutadiene	ND	250	130	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorocyclopentadiene	ND	250	110	ug/Kg	1	12/14/18	WB	SW8270D
Hexachloroethane	ND	180	110	ug/Kg	1	12/14/18	WB	SW8270D
Indeno(1,2,3-cd)pyrene	360	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Isophorone	ND	180	100	ug/Kg	1	12/14/18	WB	SW8270D
Naphthalene	ND	250	100	ug/Kg	1	12/14/18	WB	SW8270D
Nitrobenzene	ND	180	130	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodimethylamine	ND	250	100	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodi-n-propylamine	ND	180	120	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodiphenylamine	ND	250	140	ug/Kg	1	12/14/18	WB	SW8270D
Pentachloronitrobenzene	ND	250	130	ug/Kg	1	12/14/18	WB	SW8270D
Pentachlorophenol	ND	220	140	ug/Kg	1	12/14/18	WB	SW8270D
Phenanthrene	1300	250	100	ug/Kg	1	12/14/18	WB	SW8270D
Phenol	ND	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Pyrene	1500	250	120	ug/Kg	1	12/14/18	WB	SW8270D
Pyridine	ND	250	89	ug/Kg	1	12/14/18	WB	SW8270D
QA/QC Surrogates								
% 2,4,6-Tribromophenol	74			%	1	12/14/18	WB	30 - 130 %
% 2-Fluorobiphenyl	59			%	1	12/14/18	WB	30 - 130 %
% 2-Fluorophenol	54			%	1	12/14/18	WB	30 - 130 %
% Nitrobenzene-d5	58			%	1	12/14/18	WB	30 - 130 %
% Phenol-d5	61			%	1	12/14/18	WB	30 - 130 %
% Terphenyl-d14	61			%	1	12/14/18	WB	30 - 130 %
Field Extraction	Completed					12/12/18		SW5035A

1

1 = This parameter is not certified by the primary accrediting authority (NY NELAC) for this matrix. NY NELAC does not offer certification for all parameters at this time.

B = Present in blank, no bias suspected.

RL/PQL=Reporting/Practical Quantitation Level (Equivalent to NELAC LOQ, Limit of Quantitation) ND=Not Detected at RL/PQL
 BRL=Below Reporting Level L=Biased Low J=Estimated Below RL LOD=Limit of Detection MDL=Method Detection Limit

QA/QC Surrogates: Surrogates are compounds (preceded with a %) added by the lab to determine analysis efficiency. Surrogate results(%) listed in the report are not "detected" compounds.

Comments:

Per 1.4.6 of EPA method 8270D, 1,2-Diphenylhydrazine is unstable and readily converts to Azobenzene. Azobenzene is used for the calibration of 1,2-Diphenylhydrazine.

Please be advised that the NY 375 soil criteria for chromium are based on hexavalent chromium and trivalent chromium.

All soils, solids and sludges are reported on a dry weight basis unless otherwise noted in the sample comments.

S - Laboratory solvent, contamination is possible.

If there are any questions regarding this data, please call Phoenix Client Services.

This report must not be reproduced except in full as defined by the attached chain of custody.



Phyllis Shiller, Laboratory Director

December 19, 2018

Reviewed and Released by: Greg Lawrence, Assistant Lab Director



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823

Analysis Report
 December 19, 2018

FOR: Attn: Mr. Charles B. Sosik, P.G.
 Environmental Business Consultants
 1808 Middle Country Rd
 Ridge NY 11961-2406

Sample Information

Matrix: SOIL
 Location Code: EBC
 Rush Request: 72 Hour
 P.O.#:

Custody Information

Collected by: DR
 Received by: SW
 Analyzed by: see "By" below

Date

12/12/18
 12/13/18

Time

9:00
 16:30

Laboratory Data

SDG ID: GCC13976
 Phoenix ID: CC13980

Project ID: 188 EAST 135TH ST BRONX NY
 Client ID: SB 5 (8-10')

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Silver	< 0.37	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Aluminum	12400	37		mg/Kg	10	12/15/18	CPP	SW6010C
Arsenic	2.64	0.74		mg/Kg	1	12/15/18	CPP	SW6010C
Barium	27.3	0.7		mg/Kg	1	12/15/18	EK	SW6010C
Beryllium	0.53	0.29		mg/Kg	1	12/15/18	CPP	SW6010C
Calcium	7670	3.7		mg/Kg	1	12/15/18	CPP	SW6010C
Cadmium	< 0.37	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Cobalt	6.10	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Chromium	17.9	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Copper	25.3	0.7		mg/kg	1	12/15/18	CPP	SW6010C
Iron	14700	37		mg/Kg	10	12/15/18	CPP	SW6010C
Mercury	0.03	0.03		mg/Kg	1	12/14/18	RS	SW7471B
Potassium	1470	7		mg/Kg	1	12/15/18	EK	SW6010C
Magnesium	6300	37		mg/Kg	10	12/15/18	CPP	SW6010C
Manganese	110	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Sodium	289	7		mg/Kg	1	12/15/18	CPP	SW6010C
Nickel	14.4	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Lead	11.7	0.7		mg/Kg	1	12/15/18	CPP	SW6010C
Antimony	< 3.7	3.7		mg/Kg	1	12/15/18	CPP	SW6010C
Selenium	< 1.5	1.5		mg/Kg	1	12/15/18	CPP	SW6010C
Thallium	< 1.5	1.5		mg/Kg	1	12/15/18	CPP	SW6010C
Vanadium	22.5	0.37		mg/Kg	1	12/15/18	CPP	SW6010C
Zinc	64.4	0.7		mg/Kg	1	12/15/18	CPP	SW6010C
Percent Solid	85			%		12/13/18	AK	SW846-%Solid
Soil Extraction for SVOA	Completed					12/13/18	JJ/CK	SW3545A
Mercury Digestion	Completed					12/14/18	EV/EV	SW7471B
Total Metals Digest	Completed					12/14/18	M/AG	SW3050B

B

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,1-Trichloroethane	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,2,2-Tetrachloroethane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
1,1,2-Trichloroethane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloroethane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloroethene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
1,1-Dichloropropene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,3-Trichlorobenzene	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,3-Trichloropropane	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,4-Trichlorobenzene	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
1,2,4-Trimethylbenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dibromo-3-chloropropane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dibromoethane	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichlorobenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichloroethane	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
1,2-Dichloropropane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
1,3,5-Trimethylbenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
1,3-Dichlorobenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
1,3-Dichloropropane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
1,4-Dichlorobenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
2,2-Dichloropropane	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
2-Chlorotoluene	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
2-Hexanone	ND	25	5.0	ug/Kg	1	12/15/18	JLI	SW8260C
2-Isopropyltoluene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
4-Chlorotoluene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
4-Methyl-2-pentanone	ND	25	5.0	ug/Kg	1	12/15/18	JLI	SW8260C
Acetone	36	S 25	5.0	ug/Kg	1	12/15/18	JLI	SW8260C
Acrylonitrile	ND	10	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Benzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Bromobenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Bromochloromethane	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Bromodichloromethane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Bromoform	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Bromomethane	ND	5.0	2.0	ug/Kg	1	12/15/18	JLI	SW8260C
Carbon Disulfide	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Carbon tetrachloride	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Chlorobenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Chloroethane	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Chloroform	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Chloromethane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
cis-1,2-Dichloroethene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
cis-1,3-Dichloropropene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Dibromochloromethane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Dibromomethane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Dichlorodifluoromethane	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Ethylbenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Hexachlorobutadiene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Isopropylbenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
m&p-Xylene	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Methyl Ethyl Ketone	ND	30	5.0	ug/Kg	1	12/15/18	JLI	SW8260C
Methyl t-butyl ether (MTBE)	ND	10	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Methylene chloride	ND	5.0	5.0	ug/Kg	1	12/15/18	JLI	SW8260C
Naphthalene	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
n-Butylbenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
n-Propylbenzene	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
o-Xylene	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
p-Isopropyltoluene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
sec-Butylbenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Styrene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
tert-Butylbenzene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Tetrachloroethene	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Tetrahydrofuran (THF)	ND	10	2.5	ug/Kg	1	12/15/18	JLI	SW8260C
Toluene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,2-Dichloroethene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,3-Dichloropropene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
trans-1,4-dichloro-2-butene	ND	10	2.5	ug/Kg	1	12/15/18	JLI	SW8260C
Trichloroethene	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Trichlorofluoromethane	ND	5.0	1.0	ug/Kg	1	12/15/18	JLI	SW8260C
Trichlorotrifluoroethane	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
Vinyl chloride	ND	5.0	0.50	ug/Kg	1	12/15/18	JLI	SW8260C
<u>QA/QC Surrogates</u>								
% 1,2-dichlorobenzene-d4	101			%	1	12/15/18	JLI	70 - 130 %
% Bromofluorobenzene	96			%	1	12/15/18	JLI	70 - 130 %
% Dibromofluoromethane	94			%	1	12/15/18	JLI	70 - 130 %
% Toluene-d8	100			%	1	12/15/18	JLI	70 - 130 %
<u>1,4-dioxane</u>								
1,4-dioxane	ND	75		ug/kg	1	12/15/18	JLI	SW8260C
<u>QA/QC Surrogates</u>								
% 1,2-dichlorobenzene-d4	101			%	1	12/15/18	JLI	70 - 130 %
% Bromofluorobenzene	96			%	1	12/15/18	JLI	70 - 130 %
% Toluene-d8	100			%	1	12/15/18	JLI	70 - 130 %
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	20		ug/Kg	1	12/15/18	JLI	SW8260C
Acrolein	ND	5.0		ug/Kg	1	12/15/18	JLI	SW8260C
Acrylonitrile	ND	20		ug/Kg	1	12/15/18	JLI	SW8260C
Tert-butyl alcohol	ND	100		ug/Kg	1	12/15/18	JLI	SW8260C
<u>Semivolatiles</u>								
1,2,4,5-Tetrachlorobenzene	ND	270	140	ug/Kg	1	12/14/18	WB	SW8270D
1,2,4-Trichlorobenzene	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
1,2-Dichlorobenzene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
1,2-Diphenylhydrazine	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
1,3-Dichlorobenzene	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
1,4-Dichlorobenzene	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
2,4,5-Trichlorophenol	ND	270	210	ug/Kg	1	12/14/18	WB	SW8270D
2,4,6-Trichlorophenol	ND	200	120	ug/Kg	1	12/14/18	WB	SW8270D

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
2,4-Dichlorophenol	ND	200	140	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dimethylphenol	ND	270	97	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dinitrophenol	ND	270	270	ug/Kg	1	12/14/18	WB	SW8270D
2,4-Dinitrotoluene	ND	200	150	ug/Kg	1	12/14/18	WB	SW8270D
2,6-Dinitrotoluene	ND	200	120	ug/Kg	1	12/14/18	WB	SW8270D
2-Chloronaphthalene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Chlorophenol	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
2-Methylnaphthalene	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
2-Methylphenol (o-cresol)	ND	270	180	ug/Kg	1	12/14/18	WB	SW8270D
2-Nitroaniline	ND	270	270	ug/Kg	1	12/14/18	WB	SW8270D
2-Nitrophenol	ND	270	250	ug/Kg	1	12/14/18	WB	SW8270D
3&4-Methylphenol (m&p-cresol)	ND	270	150	ug/Kg	1	12/14/18	WB	SW8270D
3,3'-Dichlorobenzidine	ND	200	180	ug/Kg	1	12/14/18	WB	SW8270D
3-Nitroaniline	ND	390	780	ug/Kg	1	12/14/18	WB	SW8270D
4,6-Dinitro-2-methylphenol	ND	230	78	ug/Kg	1	12/14/18	WB	SW8270D
4-Bromophenyl phenyl ether	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
4-Chloro-3-methylphenol	ND	270	140	ug/Kg	1	12/14/18	WB	SW8270D
4-Chloroaniline	ND	310	180	ug/Kg	1	12/14/18	WB	SW8270D
4-Chlorophenyl phenyl ether	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Nitroaniline	ND	390	130	ug/Kg	1	12/14/18	WB	SW8270D
4-Nitrophenol	ND	390	180	ug/Kg	1	12/14/18	WB	SW8270D
Acenaphthene	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Acenaphthylene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Acetophenone	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Aniline	ND	310	310	ug/Kg	1	12/14/18	WB	SW8270D
Anthracene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Benz(a)anthracene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzidine	ND	390	230	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(a)pyrene	ND	200	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(b)fluoranthene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(ghi)perylene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzo(k)fluoranthene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Benzoic acid	ND	2000	780	ug/Kg	1	12/14/18	WB	SW8270D
Benzyl butyl phthalate	ND	270	100	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroethoxy)methane	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroethyl)ether	ND	200	110	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-chloroisopropyl)ether	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Bis(2-ethylhexyl)phthalate	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Carbazole	ND	200	160	ug/Kg	1	12/14/18	WB	SW8270D
Chrysene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Dibenz(a,h)anthracene	ND	200	130	ug/Kg	1	12/14/18	WB	SW8270D
Dibenzofuran	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Diethyl phthalate	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Dimethylphthalate	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Di-n-butylphthalate	ND	270	100	ug/Kg	1	12/14/18	WB	SW8270D
Di-n-octylphthalate	ND	270	100	ug/Kg	1	12/14/18	WB	SW8270D
Fluoranthene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Fluorene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorobenzene	ND	200	110	ug/Kg	1	12/14/18	WB	SW8270D

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Hexachlorobutadiene	ND	270	140	ug/Kg	1	12/14/18	WB	SW8270D
Hexachlorocyclopentadiene	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Hexachloroethane	ND	200	120	ug/Kg	1	12/14/18	WB	SW8270D
Indeno(1,2,3-cd)pyrene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Isophorone	ND	200	110	ug/Kg	1	12/14/18	WB	SW8270D
Naphthalene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Nitrobenzene	ND	200	140	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodimethylamine	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodi-n-propylamine	ND	200	130	ug/Kg	1	12/14/18	WB	SW8270D
N-Nitrosodiphenylamine	ND	270	150	ug/Kg	1	12/14/18	WB	SW8270D
Pentachloronitrobenzene	ND	270	150	ug/Kg	1	12/14/18	WB	SW8270D
Pentachlorophenol	ND	230	150	ug/Kg	1	12/14/18	WB	SW8270D
Phenanthrene	ND	270	110	ug/Kg	1	12/14/18	WB	SW8270D
Phenol	ND	270	120	ug/Kg	1	12/14/18	WB	SW8270D
Pyrene	ND	270	130	ug/Kg	1	12/14/18	WB	SW8270D
Pyridine	ND	270	96	ug/Kg	1	12/14/18	WB	SW8270D
QA/QC Surrogates								
% 2,4,6-Tribromophenol	77			%	1	12/14/18	WB	30 - 130 %
% 2-Fluorobiphenyl	62			%	1	12/14/18	WB	30 - 130 %
% 2-Fluorophenol	62			%	1	12/14/18	WB	30 - 130 %
% Nitrobenzene-d5	61			%	1	12/14/18	WB	30 - 130 %
% Phenol-d5	68			%	1	12/14/18	WB	30 - 130 %
% Terphenyl-d14	64			%	1	12/14/18	WB	30 - 130 %
Field Extraction	Completed					12/12/18		SW5035A

1

1 = This parameter is not certified by the primary accrediting authority (NY NELAC) for this matrix. NY NELAC does not offer certification for all parameters at this time.

B = Present in blank, no bias suspected.

RL/PQL=Reporting/Practical Quantitation Level (Equivalent to NELAC LOQ, Limit of Quantitation) ND=Not Detected at RL/PQL
 BRL=Below Reporting Level L=Biased Low LOD=Limit of Detection MDL=Method Detection Limit1

QA/QC Surrogates: Surrogates are compounds (preceded with a %) added by the lab to determine analysis efficiency. Surrogate results(%) listed in the report are not "detected" compounds.

Comments:

Per 1.4.6 of EPA method 8270D, 1,2-Diphenylhydrazine is unstable and readily converts to Azobenzene. Azobenzene is used for the calibration of 1,2-Diphenylhydrazine.

Please be advised that the NY 375 soil criteria for chromium are based on hexavalent chromium and trivalent chromium.

All soils, solids and sludges are reported on a dry weight basis unless otherwise noted in the sample comments.

S - Laboratory solvent, contamination is possible.

If there are any questions regarding this data, please call Phoenix Client Services.

This report must not be reproduced except in full as defined by the attached chain of custody.



Phyllis Shiller, Laboratory Director

December 19, 2018

Reviewed and Released by: Greg Lawrence, Assistant Lab Director



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823

Analysis Report
 December 19, 2018

FOR: Attn: Mr. Charles B. Sosik, P.G.
 Environmental Business Consultants
 1808 Middle Country Rd
 Ridge NY 11961-2406

Sample Information

Matrix: GROUND WATER
 Location Code: EBC
 Rush Request: 72 Hour
 P.O.#:

Custody Information

Collected by: DR
 Received by: SW
 Analyzed by: see "By" below

Date

12/12/18
 12/13/18

Time

8:40
 16:30

Laboratory Data

SDG ID: GCC13976
 Phoenix ID: CC13981

Project ID: 188 EAST 135TH ST BRONX NY
 Client ID: GW1

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1,1-Trichloroethane	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1,2,2-Tetrachloroethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1,2-Trichloroethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1-Dichloroethane	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1-Dichloroethene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1-Dichloropropene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2,3-Trichlorobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2,3-Trichloropropane	ND	0.25	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2,4-Trichlorobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2,4-Trimethylbenzene	1.8	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2-Dibromo-3-chloropropane	ND	0.50	0.50	ug/L	1	12/15/18	MH	SW8260C
1,2-Dibromoethane	ND	0.25	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2-Dichlorobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2-Dichloroethane	ND	0.60	0.50	ug/L	1	12/15/18	MH	SW8260C
1,2-Dichloropropane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,3,5-Trimethylbenzene	0.34	J 1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,3-Dichlorobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,3-Dichloropropane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,4-Dichlorobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
2,2-Dichloropropane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
2-Chlorotoluene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
2-Hexanone	ND	2.5	2.5	ug/L	1	12/15/18	MH	SW8260C
2-Isopropyltoluene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
4-Chlorotoluene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
4-Methyl-2-pentanone	ND	2.5	2.5	ug/L	1	12/15/18	MH	SW8260C

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Acetone	3.9	JS 5.0	2.5	ug/L	1	12/15/18	MH	SW8260C
Acrolein	ND	5.0	2.5	ug/L	1	12/15/18	MH	SW8260C
Acrylonitrile	ND	5.0	2.5	ug/L	1	12/15/18	MH	SW8260C
Benzene	ND	0.70	0.25	ug/L	1	12/15/18	MH	SW8260C
Bromobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Bromochloromethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Bromodichloromethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Bromoform	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Bromomethane	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Carbon Disulfide	0.40	J 1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Carbon tetrachloride	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Chlorobenzene	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Chloroethane	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Chloroform	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Chloromethane	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
cis-1,2-Dichloroethene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
cis-1,3-Dichloropropene	ND	0.40	0.25	ug/L	1	12/15/18	MH	SW8260C
Dibromochloromethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Dibromomethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Dichlorodifluoromethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Ethylbenzene	0.32	J 1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Hexachlorobutadiene	ND	0.50	0.20	ug/L	1	12/15/18	MH	SW8260C
Isopropylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
m&p-Xylene	1.2	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Methyl ethyl ketone	ND	2.5	2.5	ug/L	1	12/15/18	MH	SW8260C
Methyl t-butyl ether (MTBE)	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Methylene chloride	ND	3.0	1.0	ug/L	1	12/15/18	MH	SW8260C
Naphthalene	1.6	1.0	1.0	ug/L	1	12/15/18	MH	SW8260C
n-Butylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
n-Propylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
o-Xylene	0.60	J 1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
p-Isopropyltoluene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
sec-Butylbenzene	1.5	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Styrene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
tert-Butylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Tetrachloroethene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Tetrahydrofuran (THF)	ND	5.0	2.5	ug/L	1	12/15/18	MH	SW8260C
Toluene	1.3	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
trans-1,2-Dichloroethene	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
trans-1,3-Dichloropropene	ND	0.40	0.25	ug/L	1	12/15/18	MH	SW8260C
trans-1,4-dichloro-2-butene	ND	2.5	2.5	ug/L	1	12/15/18	MH	SW8260C
Trichloroethene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Trichlorofluoromethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Trichlorotrifluoroethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Vinyl chloride	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
QA/QC Surrogates								
% 1,2-dichlorobenzene-d4	101			%	1	12/15/18	MH	70 - 130 %
% Bromofluorobenzene	96			%	1	12/15/18	MH	70 - 130 %
% Dibromofluoromethane	100			%	1	12/15/18	MH	70 - 130 %

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
% Toluene-d8	98			%	1	12/15/18	MH	70 - 130 %
<u>1,4-dioxane</u>								
1,4-dioxane	ND	100		ug/l	1	12/15/18	MH	SW8260C
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	1.0		ug/L	1	12/15/18	MH	SW8260C
Acrolein	ND	5.0		ug/L	1	12/15/18	MH	SW8260C
Acrylonitrile	ND	5.0		ug/L	1	12/15/18	MH	SW8260C
Tert-butyl alcohol	ND	50		ug/L	1	12/15/18	MH	SW8260C

1 = This parameter is not certified by the primary accrediting authority (NY NELAC) for this matrix. NY NELAC does not offer certification for all parameters at this time.

RL/PQL=Reporting/Practical Quantitation Level (Equivalent to NELAC LOQ, Limit of Quantitation) ND=Not Detected at RL/PQL
BRL=Below Reporting Level L=Biased Low J=Estimated Below RL LOD=Limit of Detection MDL=Method Detection Limit
QA/QC Surrogates: Surrogates are compounds (preceded with a %) added by the lab to determine analysis efficiency. Surrogate results(%) listed in the report are not "detected" compounds.


Comments:

Volatile Comment:

Where the LOD justifies lowering the RL/PQL, the RL/PQL of some compounds are evaluated below the lowest calibration standard in order to meet criteria.

S - Laboratory solvent, contamination is possible.

If there are any questions regarding this data, please call Phoenix Client Services.
This report must not be reproduced except in full as defined by the attached chain of custody.



Phyllis Shiller, Laboratory Director

December 19, 2018

Reviewed and Released by: Greg Lawrence, Assistant Lab Director



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823

Analysis Report
 December 19, 2018

FOR: Attn: Mr. Charles B. Sosik, P.G.
 Environmental Business Consultants
 1808 Middle Country Rd
 Ridge NY 11961-2406

Sample Information

Matrix: SOIL
 Location Code: EBC
 Rush Request: 72 Hour
 P.O.#:

Custody Information

Collected by: DR
 Received by: SW
 Analyzed by: see "By" below

Date

12/12/18
 12/13/18

Time

10:00
 16:30

Laboratory Data

SDG ID: GCC13976
 Phoenix ID: CC13982

Project ID: 188 EAST 135TH ST BRONX NY
 Client ID: GW2

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,1,1-Trichloroethane	ND	5.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,1,2,2-Tetrachloroethane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,1,2-Trichloroethane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,1-Dichloroethane	ND	5.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,1-Dichloroethene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,1-Dichloropropene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,2,3-Trichlorobenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,2,3-Trichloropropane	ND	0.25	0.25	ug/L	1	12/14/18	MH	SW8260C
1,2,4-Trichlorobenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,2,4-Trimethylbenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,2-Dibromo-3-chloropropane	ND	0.50	0.50	ug/L	1	12/14/18	MH	SW8260C
1,2-Dibromoethane	ND	0.25	0.25	ug/L	1	12/14/18	MH	SW8260C
1,2-Dichlorobenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,2-Dichloroethane	ND	0.60	0.50	ug/L	1	12/14/18	MH	SW8260C
1,2-Dichloropropane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,3,5-Trimethylbenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,3-Dichlorobenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,3-Dichloropropane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
1,4-Dichlorobenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
2,2-Dichloropropane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
2-Chlorotoluene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
2-Hexanone	ND	2.5	2.5	ug/L	1	12/14/18	MH	SW8260C
2-Isopropyltoluene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
4-Chlorotoluene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
4-Methyl-2-pentanone	ND	2.5	2.5	ug/L	1	12/14/18	MH	SW8260C

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Acetone	ND	5.0	2.5	ug/L	1	12/14/18	MH	SW8260C
Acrolein	ND	5.0	2.5	ug/L	1	12/14/18	MH	SW8260C
Acrylonitrile	ND	5.0	2.5	ug/L	1	12/14/18	MH	SW8260C
Benzene	ND	0.70	0.25	ug/L	1	12/14/18	MH	SW8260C
Bromobenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Bromochloromethane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Bromodichloromethane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Bromoform	ND	5.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Bromomethane	ND	5.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Carbon Disulfide	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Carbon tetrachloride	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Chlorobenzene	ND	5.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Chloroethane	ND	5.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Chloroform	ND	5.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Chloromethane	ND	5.0	0.25	ug/L	1	12/14/18	MH	SW8260C
cis-1,2-Dichloroethene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
cis-1,3-Dichloropropene	ND	0.40	0.25	ug/L	1	12/14/18	MH	SW8260C
Dibromochloromethane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Dibromomethane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Dichlorodifluoromethane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Ethylbenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Hexachlorobutadiene	ND	0.50	0.20	ug/L	1	12/14/18	MH	SW8260C
Isopropylbenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
m&p-Xylene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Methyl ethyl ketone	ND	2.5	2.5	ug/L	1	12/14/18	MH	SW8260C
Methyl t-butyl ether (MTBE)	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Methylene chloride	ND	3.0	1.0	ug/L	1	12/14/18	MH	SW8260C
Naphthalene	ND	1.0	1.0	ug/L	1	12/14/18	MH	SW8260C
n-Butylbenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
n-Propylbenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
o-Xylene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
p-Isopropyltoluene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
sec-Butylbenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Styrene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
tert-Butylbenzene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Tetrachloroethene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Tetrahydrofuran (THF)	ND	5.0	2.5	ug/L	1	12/14/18	MH	SW8260C
Toluene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
trans-1,2-Dichloroethene	ND	5.0	0.25	ug/L	1	12/14/18	MH	SW8260C
trans-1,3-Dichloropropene	ND	0.40	0.25	ug/L	1	12/14/18	MH	SW8260C
trans-1,4-dichloro-2-butene	ND	2.5	2.5	ug/L	1	12/14/18	MH	SW8260C
Trichloroethene	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Trichlorofluoromethane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Trichlorotrifluoroethane	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
Vinyl chloride	ND	1.0	0.25	ug/L	1	12/14/18	MH	SW8260C
QA/QC Surrogates								
% 1,2-dichlorobenzene-d4	99			%	1	12/14/18	MH	70 - 130 %
% Bromofluorobenzene	97			%	1	12/14/18	MH	70 - 130 %
% Dibromofluoromethane	98			%	1	12/14/18	MH	70 - 130 %

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
% Toluene-d8	99			%	1	12/14/18	MH	70 - 130 %
<u>1,4-dioxane</u>								
1,4-dioxane	ND	100		ug/l	1	12/14/18	MH	SW8260C
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	1.0		ug/L	1	12/14/18	MH	SW8260C
Acrolein	ND	5.0		ug/L	1	12/14/18	MH	SW8260C
Acrylonitrile	ND	5.0		ug/L	1	12/14/18	MH	SW8260C
Tert-butyl alcohol	ND	50		ug/L	1	12/14/18	MH	SW8260C

1 = This parameter is not certified by the primary accrediting authority (NY NELAC) for this matrix. NY NELAC does not offer certification for all parameters at this time.

RL/PQL=Reporting/Practical Quantitation Level (Equivalent to NELAC LOQ, Limit of Quantitation) ND=Not Detected at RL/PQL
BRL=Below Reporting Level L=Biased Low LOD=Limit of Detection MDL=Method Detection Limit

QA/QC Surrogates: Surrogates are compounds (preceded with a %) added by the lab to determine analysis efficiency. Surrogate results(%) listed in the report are not "detected" compounds.


Comments:

Volatile Comment:

Where the LOD justifies lowering the RL/PQL, the RL/PQL of some compounds are evaluated below the lowest calibration standard in order to meet criteria.

If there are any questions regarding this data, please call Phoenix Client Services.

This report must not be reproduced except in full as defined by the attached chain of custody.



Phyllis Shiller, Laboratory Director

December 19, 2018

Reviewed and Released by: Greg Lawrence, Assistant Lab Director



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823

Analysis Report
 December 19, 2018

FOR: Attn: Mr. Charles B. Sosik, P.G.
 Environmental Business Consultants
 1808 Middle Country Rd
 Ridge NY 11961-2406

Sample Information

Matrix: SOIL
 Location Code: EBC
 Rush Request: 72 Hour
 P.O.#:

Custody Information

Collected by: DR
 Received by: SW
 Analyzed by: see "By" below

Date

12/12/18
 12/13/18

Time

10:45
 16:30

Laboratory Data

SDG ID: GCC13976
 Phoenix ID: CC13983

Project ID: 188 EAST 135TH ST BRONX NY
 Client ID: GW3

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1,1-Trichloroethane	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1,2,2-Tetrachloroethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1,2-Trichloroethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1-Dichloroethane	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1-Dichloroethene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,1-Dichloropropene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2,3-Trichlorobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2,3-Trichloropropane	ND	0.25	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2,4-Trichlorobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2,4-Trimethylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2-Dibromo-3-chloropropane	ND	0.50	0.50	ug/L	1	12/15/18	MH	SW8260C
1,2-Dibromoethane	ND	0.25	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2-Dichlorobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,2-Dichloroethane	ND	0.60	0.50	ug/L	1	12/15/18	MH	SW8260C
1,2-Dichloropropane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,3,5-Trimethylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,3-Dichlorobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,3-Dichloropropane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
1,4-Dichlorobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
2,2-Dichloropropane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
2-Chlorotoluene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
2-Hexanone	ND	2.5	2.5	ug/L	1	12/15/18	MH	SW8260C
2-Isopropyltoluene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
4-Chlorotoluene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
4-Methyl-2-pentanone	ND	2.5	2.5	ug/L	1	12/15/18	MH	SW8260C

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
Acetone	3.7	JS 5.0	2.5	ug/L	1	12/15/18	MH	SW8260C
Acrolein	ND	5.0	2.5	ug/L	1	12/15/18	MH	SW8260C
Acrylonitrile	ND	5.0	2.5	ug/L	1	12/15/18	MH	SW8260C
Benzene	ND	0.70	0.25	ug/L	1	12/15/18	MH	SW8260C
Bromobenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Bromochloromethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Bromodichloromethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Bromoform	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Bromomethane	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Carbon Disulfide	0.39	J 1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Carbon tetrachloride	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Chlorobenzene	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Chloroethane	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Chloroform	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Chloromethane	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
cis-1,2-Dichloroethene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
cis-1,3-Dichloropropene	ND	0.40	0.25	ug/L	1	12/15/18	MH	SW8260C
Dibromochloromethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Dibromomethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Dichlorodifluoromethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Ethylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Hexachlorobutadiene	ND	0.50	0.20	ug/L	1	12/15/18	MH	SW8260C
Isopropylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
m&p-Xylene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Methyl ethyl ketone	ND	2.5	2.5	ug/L	1	12/15/18	MH	SW8260C
Methyl t-butyl ether (MTBE)	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Methylene chloride	ND	3.0	1.0	ug/L	1	12/15/18	MH	SW8260C
Naphthalene	ND	1.0	1.0	ug/L	1	12/15/18	MH	SW8260C
n-Butylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
n-Propylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
o-Xylene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
p-Isopropyltoluene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
sec-Butylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Styrene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
tert-Butylbenzene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Tetrachloroethene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Tetrahydrofuran (THF)	ND	5.0	2.5	ug/L	1	12/15/18	MH	SW8260C
Toluene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
trans-1,2-Dichloroethene	ND	5.0	0.25	ug/L	1	12/15/18	MH	SW8260C
trans-1,3-Dichloropropene	ND	0.40	0.25	ug/L	1	12/15/18	MH	SW8260C
trans-1,4-dichloro-2-butene	ND	2.5	2.5	ug/L	1	12/15/18	MH	SW8260C
Trichloroethene	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Trichlorofluoromethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Trichlorotrifluoroethane	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
Vinyl chloride	ND	1.0	0.25	ug/L	1	12/15/18	MH	SW8260C
QA/QC Surrogates								
% 1,2-dichlorobenzene-d4	100			%	1	12/15/18	MH	70 - 130 %
% Bromofluorobenzene	96			%	1	12/15/18	MH	70 - 130 %
% Dibromofluoromethane	100			%	1	12/15/18	MH	70 - 130 %

Parameter	Result	RL/ PQL	LOD/ MDL	Units	Dilution	Date/Time	By	Reference
% Toluene-d8	99			%	1	12/15/18	MH	70 - 130 %
<u>1,4-dioxane</u>								
1,4-dioxane	ND	100		ug/l	1	12/15/18	MH	SW8260C
<u>Volatiles</u>								
1,1,1,2-Tetrachloroethane	ND	1.0		ug/L	1	12/15/18	MH	SW8260C
Acrolein	ND	5.0		ug/L	1	12/15/18	MH	SW8260C
Acrylonitrile	ND	5.0		ug/L	1	12/15/18	MH	SW8260C
Tert-butyl alcohol	ND	50		ug/L	1	12/15/18	MH	SW8260C

1 = This parameter is not certified by the primary accrediting authority (NY NELAC) for this matrix. NY NELAC does not offer certification for all parameters at this time.

RL/PQL=Reporting/Practical Quantitation Level (Equivalent to NELAC LOQ, Limit of Quantitation) ND=Not Detected at RL/PQL
BRL=Below Reporting Level L=Biased Low J=Estimated Below RL LOD=Limit of Detection MDL=Method Detection Limit
QA/QC Surrogates: Surrogates are compounds (preceded with a %) added by the lab to determine analysis efficiency. Surrogate results(%) listed in the report are not "detected" compounds.


Comments:

Volatile Comment:

Where the LOD justifies lowering the RL/PQL, the RL/PQL of some compounds are evaluated below the lowest calibration standard in order to meet criteria.

S - Laboratory solvent, contamination is possible.

If there are any questions regarding this data, please call Phoenix Client Services.
This report must not be reproduced except in full as defined by the attached chain of custody.



Phyllis Shiller, Laboratory Director

December 19, 2018

Reviewed and Released by: Greg Lawrence, Assistant Lab Director



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823



QA/QC Report

December 19, 2018

QA/QC Data

SDG I.D.: GCC13976

Parameter	Blank	Blk RL	Sample Result	Dup Result	Dup RPD	LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
-----------	-------	--------	---------------	------------	---------	-------	--------	---------	------	-------	--------	--------------	--------------

QA/QC Batch 459938 (mg/kg), QC Sample No: CC12609 (CC13976, CC13977, CC13978, CC13979, CC13980)

Mercury - Soil	BRL	0.02	0.20	0.29	NC	82.0	91.7	11.2	59.3	61.2	3.2	70 - 130	30 m
----------------	-----	------	------	------	----	------	------	------	------	------	-----	----------	------

Comment: Additional Mercury criteria: LCS acceptance range for waters is 80-120% and for soils is 70-130%. MS acceptance range is 75-125%.

QA/QC Batch 460033 (mg/kg), QC Sample No: CC14009 (CC13976, CC13977, CC13978, CC13979, CC13980)

ICP Metals - Soil

Aluminum	BRL	5.1	9510	9690	1.90	110				NC		75 - 125	30
Antimony	BRL	3.4	<3.5	<3.4	NC	123				88.6		75 - 125	30
Arsenic	BRL	0.68	7.18	7.57	5.30	101				99.8		75 - 125	30
Barium	0.90	0.34	50.8	62.5	20.7	99.1				110		75 - 125	30
Beryllium	BRL	0.27	0.44	0.47	NC	96.7				101		75 - 125	30
Cadmium	BRL	0.34	<0.35	0.34	NC	98.3				101		75 - 125	30
Calcium	BRL	5.1	1510	1590	5.20	97.8				NC		75 - 125	30
Chromium	BRL	0.34	24.7	29.2	16.7	102				108		75 - 125	30
Cobalt	BRL	0.34	2.98	3.08	3.30	103				103		75 - 125	30
Copper	BRL	0.68	4.8	5.01	4.30	95.5				102		75 - 125	30
Iron	BRL	5.1	20000	20500	2.50	125				NC		75 - 125	30
Lead	BRL	0.34	5.6	5.82	3.90	101				102		75 - 125	30
Magnesium	BRL	5.1	1480	1570	5.90	106				NC		75 - 125	30
Manganese	BRL	0.34	62.7	63.9	1.90	101				107		75 - 125	30
Nickel	BRL	0.34	4.75	5.02	5.50	97.3				103		75 - 125	30
Potassium	BRL	5.1	1460	1580	7.90	113				NC		75 - 125	30
Selenium	BRL	1.4	<1.4	<1.4	NC	99.7				99.9		75 - 125	30
Silver	BRL	0.34	<0.35	<0.34	NC	95.3				101		75 - 125	30
Sodium	BRL	5.1	51	123	82.8	99.0				111		75 - 125	30 r
Thallium	BRL	3.1	<1.4	<3.1	NC	98.8				102		75 - 125	30
Vanadium	BRL	0.34	28.2	29.6	4.80	114				106		75 - 125	30
Zinc	BRL	0.68	23.0	25.5	10.3	97.8				101		75 - 125	30

m = This parameter is outside laboratory MS/MSD specified recovery limits.

r = This parameter is outside laboratory RPD specified recovery limits.



Environmental Laboratories, Inc.
 587 East Middle Turnpike, P.O. Box 370, Manchester, CT 06045
 Tel. (860) 645-1102 Fax (860) 645-0823



QA/QC Report

December 19, 2018

QA/QC Data

SDG I.D.: GCC13976

Parameter	Blank	Blk RL	LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
QA/QC Batch 459811 (ug/kg), QC Sample No: CC13734 (CC13976, CC13977, CC13978, CC13979, CC13980)										
Semivolatiles - Soil										
1,2,4,5-Tetrachlorobenzene	ND	230	73	70	4.2	73	69	5.6	30 - 130	30
1,2,4-Trichlorobenzene	ND	230	71	69	2.9	68	64	6.1	30 - 130	30
1,2-Dichlorobenzene	ND	180	65	63	3.1	61	58	5.0	30 - 130	30
1,2-Diphenylhydrazine	ND	230	68	66	3.0	70	66	5.9	30 - 130	30
1,3-Dichlorobenzene	ND	230	64	62	3.2	55	53	3.7	30 - 130	30
1,4-Dichlorobenzene	ND	230	65	63	3.1	59	56	5.2	30 - 130	30
2,4,5-Trichlorophenol	ND	230	84	77	8.7	86	80	7.2	30 - 130	30
2,4,6-Trichlorophenol	ND	130	82	78	5.0	84	78	7.4	30 - 130	30
2,4-Dichlorophenol	ND	130	80	78	2.5	80	77	3.8	30 - 130	30
2,4-Dimethylphenol	ND	230	81	79	2.5	85	81	4.8	30 - 130	30
2,4-Dinitrophenol	ND	230	14	<10	NC	31	34	9.2	30 - 130	30
2,4-Dinitrotoluene	ND	130	79	78	1.3	83	81	2.4	30 - 130	30
2,6-Dinitrotoluene	ND	130	81	80	1.2	84	82	2.4	30 - 130	30
2-Chloronaphthalene	ND	230	75	72	4.1	76	69	9.7	30 - 130	30
2-Chlorophenol	ND	230	72	71	1.4	76	70	8.2	30 - 130	30
2-Methylnaphthalene	ND	230	72	70	2.8	73	69	5.6	30 - 130	30
2-Methylphenol (o-cresol)	ND	230	70	69	1.4	76	72	5.4	30 - 130	30
2-Nitroaniline	ND	330	103	96	7.0	101	101	0.0	30 - 130	30
2-Nitrophenol	ND	230	75	76	1.3	76	73	4.0	30 - 130	30
3&4-Methylphenol (m&p-cresol)	ND	230	76	75	1.3	83	79	4.9	30 - 130	30
3,3'-Dichlorobenzidine	ND	130	78	80	2.5	96	90	6.5	30 - 130	30
3-Nitroaniline	ND	330	93	90	3.3	97	96	1.0	30 - 130	30
4,6-Dinitro-2-methylphenol	ND	230	35	22	45.6	44	47	6.6	30 - 130	30
4-Bromophenyl phenyl ether	ND	230	82	79	3.7	82	73	11.6	30 - 130	30
4-Chloro-3-methylphenol	ND	230	81	79	2.5	82	82	0.0	30 - 130	30
4-Chloroaniline	ND	230	74	74	0.0	81	78	3.8	30 - 130	30
4-Chlorophenyl phenyl ether	ND	230	81	77	5.1	79	74	6.5	30 - 130	30
4-Nitroaniline	ND	230	79	77	2.6	83	82	1.2	30 - 130	30
4-Nitrophenol	ND	230	76	74	2.7	79	77	2.6	30 - 130	30
Acenaphthene	ND	230	76	72	5.4	74	70	5.6	30 - 130	30
Acenaphthylene	ND	130	70	68	2.9	71	66	7.3	30 - 130	30
Acetophenone	ND	230	62	61	1.6	70	65	7.4	30 - 130	30
Aniline	ND	330	58	58	0.0	110	60	58.8	30 - 130	30
Anthracene	ND	230	76	75	1.3	79	72	9.3	30 - 130	30
Benz(a)anthracene	ND	230	75	74	1.3	73	67	8.6	30 - 130	30
Benzidine	ND	330	55	55	0.0	45	61	30.2	30 - 130	30
Benzo(a)pyrene	ND	130	75	73	2.7	73	66	10.1	30 - 130	30
Benzo(b)fluoranthene	ND	160	79	77	2.6	72	68	5.7	30 - 130	30
Benzo(ghi)perylene	ND	230	72	71	1.4	76	68	11.1	30 - 130	30
Benzo(k)fluoranthene	ND	230	76	72	5.4	71	65	8.8	30 - 130	30
Benzoic Acid	ND	330	<10	<10	NC	35	35	0.0	30 - 130	30

QA/QC Data

SDG I.D.: GCC13976

Parameter	Blk		LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
	Blank	RL								
Benzyl butyl phthalate	ND	230	84	81	3.6	81	73	10.4	30 - 130	30
Bis(2-chloroethoxy)methane	ND	230	71	71	0.0	70	66	5.9	30 - 130	30
Bis(2-chloroethyl)ether	ND	130	60	58	3.4	61	57	6.8	30 - 130	30
Bis(2-chloroisopropyl)ether	ND	230	53	52	1.9	52	49	5.9	30 - 130	30
Bis(2-ethylhexyl)phthalate	ND	230	86	83	3.6	90	78	14.3	30 - 130	30
Carbazole	ND	230	77	75	2.6	81	74	9.0	30 - 130	30
Chrysene	ND	230	78	75	3.9	78	71	9.4	30 - 130	30
Dibenz(a,h)anthracene	ND	130	80	79	1.3	85	75	12.5	30 - 130	30
Dibenzofuran	ND	230	76	72	5.4	77	72	6.7	30 - 130	30
Diethyl phthalate	ND	230	80	77	3.8	79	75	5.2	30 - 130	30
Dimethylphthalate	ND	230	80	77	3.8	80	75	6.5	30 - 130	30
Di-n-butylphthalate	ND	670	84	82	2.4	83	76	8.8	30 - 130	30
Di-n-octylphthalate	ND	230	87	84	3.5	86	78	9.8	30 - 130	30
Fluoranthene	ND	230	79	76	3.9	77	72	6.7	30 - 130	30
Fluorene	ND	230	78	74	5.3	77	73	5.3	30 - 130	30
Hexachlorobenzene	ND	130	75	73	2.7	73	64	13.1	30 - 130	30
Hexachlorobutadiene	ND	230	73	72	1.4	66	63	4.7	30 - 130	30
Hexachlorocyclopentadiene	ND	230	53	53	0.0	22	12	58.8	30 - 130	30
Hexachloroethane	ND	130	63	61	3.2	53	50	5.8	30 - 130	30
Indeno(1,2,3-cd)pyrene	ND	230	71	72	1.4	76	68	11.1	30 - 130	30
Isophorone	ND	130	66	65	1.5	65	62	4.7	30 - 130	30
Naphthalene	ND	230	69	68	1.5	69	65	6.0	30 - 130	30
Nitrobenzene	ND	130	68	68	0.0	72	67	7.2	30 - 130	30
N-Nitrosodimethylamine	ND	230	59	57	3.4	63	58	8.3	30 - 130	30
N-Nitrosodi-n-propylamine	ND	130	70	69	1.4	73	68	7.1	30 - 130	30
N-Nitrosodiphenylamine	ND	130	80	77	3.8	79	75	5.2	30 - 130	30
Pentachloronitrobenzene	ND	230	77	76	1.3	79	72	9.3	30 - 130	30
Pentachlorophenol	ND	230	78	71	9.4	86	81	6.0	30 - 130	30
Phenanthrene	ND	130	76	73	4.0	77	73	5.3	30 - 130	30
Phenol	ND	230	72	70	2.8	75	72	4.1	30 - 130	30
Pyrene	ND	230	79	77	2.6	79	74	6.5	30 - 130	30
Pyridine	ND	230	44	42	4.7	44	42	4.7	30 - 130	30
% 2,4,6-Tribromophenol	61	%	74	71	4.1	79	72	9.3	30 - 130	30
% 2-Fluorobiphenyl	58	%	62	61	1.6	64	59	8.1	30 - 130	30
% 2-Fluorophenol	55	%	66	64	3.1	72	65	10.2	30 - 130	30
% Nitrobenzene-d5	53	%	58	58	0.0	63	59	6.6	30 - 130	30
% Phenol-d5	61	%	70	68	2.9	74	71	4.1	30 - 130	30
% Terphenyl-d14	54	%	62	60	3.3	63	57	10.0	30 - 130	30

m,r

Comment:

Additional 8270 criteria: 20% of compounds can be outside of acceptance criteria as long as recovery is at least 10%. (Acid surrogates acceptance range for aqueous samples: 15-110%, for soils 30-130%)

QA/QC Batch 460180 (ug/L), QC Sample No: CC12929 (CC13981, CC13983)

Volatiles - Ground Water

1,1,1,2-Tetrachloroethane	ND	1.0	109	109	0.0				70 - 130	30
1,1,1-Trichloroethane	ND	1.0	100	100	0.0				70 - 130	30
1,1,2,2-Tetrachloroethane	ND	0.50	111	108	2.7				70 - 130	30
1,1,2-Trichloroethane	ND	1.0	107	110	2.8				70 - 130	30
1,1-Dichloroethane	ND	1.0	98	99	1.0				70 - 130	30
1,1-Dichloroethene	ND	1.0	100	101	1.0				70 - 130	30
1,1-Dichloropropene	ND	1.0	103	101	2.0				70 - 130	30
1,2,3-Trichlorobenzene	ND	1.0	115	115	0.0				70 - 130	30
1,2,3-Trichloropropane	ND	1.0	105	103	1.9				70 - 130	30

QA/QC Data

SDG I.D.: GCC13976

Parameter	Blank	Blk RL	LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
1,2,4-Trichlorobenzene	ND	1.0	108	110	1.8				70 - 130	30
1,2,4-Trimethylbenzene	ND	1.0	99	100	1.0				70 - 130	30
1,2-Dibromo-3-chloropropane	ND	1.0	118	124	5.0				70 - 130	30
1,2-Dibromoethane	ND	1.0	111	108	2.7				70 - 130	30
1,2-Dichlorobenzene	ND	1.0	105	106	0.9				70 - 130	30
1,2-Dichloroethane	ND	1.0	106	106	0.0				70 - 130	30
1,2-Dichloropropane	ND	1.0	103	101	2.0				70 - 130	30
1,3,5-Trimethylbenzene	ND	1.0	98	99	1.0				70 - 130	30
1,3-Dichlorobenzene	ND	1.0	101	102	1.0				70 - 130	30
1,3-Dichloropropane	ND	1.0	108	106	1.9				70 - 130	30
1,4-Dichlorobenzene	ND	1.0	102	103	1.0				70 - 130	30
1,4-dioxane	ND	100	109	110	0.9				70 - 130	30
2,2-Dichloropropane	ND	1.0	106	105	0.9				70 - 130	30
2-Chlorotoluene	ND	1.0	98	100	2.0				70 - 130	30
2-Hexanone	ND	5.0	87	86	1.2				70 - 130	30
2-Isopropyltoluene	ND	1.0	101	103	2.0				70 - 130	30
4-Chlorotoluene	ND	1.0	98	99	1.0				70 - 130	30
4-Methyl-2-pentanone	ND	5.0	95	93	2.1				70 - 130	30
Acetone	ND	5.0	75	75	0.0				70 - 130	30
Acrolein	ND	5.0	110	110	0.0				70 - 130	30
Acrylonitrile	ND	5.0	108	104	3.8				70 - 130	30
Benzene	ND	0.70	100	98	2.0				70 - 130	30
Bromobenzene	ND	1.0	102	101	1.0				70 - 130	30
Bromochloromethane	ND	1.0	104	102	1.9				70 - 130	30
Bromodichloromethane	ND	0.50	111	108	2.7				70 - 130	30
Bromoform	ND	1.0	119	112	6.1				70 - 130	30
Bromomethane	ND	1.0	79	81	2.5				70 - 130	30
Carbon Disulfide	ND	1.0	99	97	2.0				70 - 130	30
Carbon tetrachloride	ND	1.0	100	99	1.0				70 - 130	30
Chlorobenzene	ND	1.0	100	101	1.0				70 - 130	30
Chloroethane	ND	1.0	103	106	2.9				70 - 130	30
Chloroform	ND	1.0	99	96	3.1				70 - 130	30
Chloromethane	ND	1.0	84	84	0.0				70 - 130	30
cis-1,2-Dichloroethene	ND	1.0	101	102	1.0				70 - 130	30
cis-1,3-Dichloropropene	ND	0.40	109	106	2.8				70 - 130	30
Dibromochloromethane	ND	0.50	117	117	0.0				70 - 130	30
Dibromomethane	ND	1.0	108	104	3.8				70 - 130	30
Dichlorodifluoromethane	ND	1.0	104	105	1.0				70 - 130	30
Ethylbenzene	ND	1.0	101	102	1.0				70 - 130	30
Hexachlorobutadiene	ND	0.40	103	107	3.8				70 - 130	30
Isopropylbenzene	ND	1.0	98	99	1.0				70 - 130	30
m&p-Xylene	ND	1.0	100	100	0.0				70 - 130	30
Methyl ethyl ketone	ND	5.0	89	89	0.0				70 - 130	30
Methyl t-butyl ether (MTBE)	ND	1.0	109	112	2.7				70 - 130	30
Methylene chloride	ND	1.0	99	98	1.0				70 - 130	30
Naphthalene	ND	1.0	119	117	1.7				70 - 130	30
n-Butylbenzene	ND	1.0	101	103	2.0				70 - 130	30
n-Propylbenzene	ND	1.0	98	99	1.0				70 - 130	30
o-Xylene	ND	1.0	103	102	1.0				70 - 130	30
p-Isopropyltoluene	ND	1.0	99	101	2.0				70 - 130	30
sec-Butylbenzene	ND	1.0	102	104	1.9				70 - 130	30
Styrene	ND	1.0	103	103	0.0				70 - 130	30
tert-butyl alcohol	ND	10	113	121	6.8				70 - 130	30

QA/QC Data

SDG I.D.: GCC13976

Parameter	Blk		LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
	Blank	RL								
tert-Butylbenzene	ND	1.0	97	99	2.0				70 - 130	30
Tetrachloroethene	ND	1.0	102	102	0.0				70 - 130	30
Tetrahydrofuran (THF)	ND	2.5	102	103	1.0				70 - 130	30
Toluene	ND	1.0	101	100	1.0				70 - 130	30
trans-1,2-Dichloroethene	ND	1.0	101	101	0.0				70 - 130	30
trans-1,3-Dichloropropene	ND	0.40	106	103	2.9				70 - 130	30
trans-1,4-dichloro-2-butene	ND	5.0	116	115	0.9				70 - 130	30
Trichloroethene	ND	1.0	103	102	1.0				70 - 130	30
Trichlorofluoromethane	ND	1.0	104	102	1.9				70 - 130	30
Trichlorotrifluoroethane	ND	1.0	99	104	4.9				70 - 130	30
Vinyl chloride	ND	1.0	102	100	2.0				70 - 130	30
% 1,2-dichlorobenzene-d4	103	%	102	101	1.0				70 - 130	30
% Bromofluorobenzene	97	%	102	102	0.0				70 - 130	30
% Dibromofluoromethane	100	%	101	100	1.0				70 - 130	30
% Toluene-d8	98	%	101	100	1.0				70 - 130	30

Comment:

A LCS and LCS Duplicate were performed instead of a matrix spike and matrix spike duplicate.

Additional 8260 criteria: 10% of LCS/LCSD compounds can be outside of acceptance criteria as long as recovery is 40-160%.

QA/QC Batch 460184 (ug/L), QC Sample No: CC13632 (CC13982)

Volatiles

1,1,1,2-Tetrachloroethane	ND	1.0	96	104	8.0				70 - 130	30
1,1,1-Trichloroethane	ND	1.0	87	97	10.9				70 - 130	30
1,1,2,2-Tetrachloroethane	ND	0.50	93	105	12.1				70 - 130	30
1,1,2-Trichloroethane	ND	1.0	93	105	12.1				70 - 130	30
1,1-Dichloroethane	ND	1.0	88	96	8.7				70 - 130	30
1,1-Dichloroethene	ND	1.0	91	99	8.4				70 - 130	30
1,1-Dichloropropene	ND	1.0	90	98	8.5				70 - 130	30
1,2,3-Trichlorobenzene	ND	1.0	93	106	13.1				70 - 130	30
1,2,3-Trichloropropane	ND	1.0	89	96	7.6				70 - 130	30
1,2,4-Trichlorobenzene	ND	1.0	92	103	11.3				70 - 130	30
1,2,4-Trimethylbenzene	ND	1.0	91	95	4.3				70 - 130	30
1,2-Dibromo-3-chloropropane	ND	1.0	98	107	8.8				70 - 130	30
1,2-Dibromoethane	ND	1.0	93	104	11.2				70 - 130	30
1,2-Dichlorobenzene	ND	1.0	93	100	7.3				70 - 130	30
1,2-Dichloroethane	ND	1.0	87	101	14.9				70 - 130	30
1,2-Dichloropropane	ND	1.0	88	97	9.7				70 - 130	30
1,3,5-Trimethylbenzene	ND	1.0	91	95	4.3				70 - 130	30
1,3-Dichlorobenzene	ND	1.0	92	98	6.3				70 - 130	30
1,3-Dichloropropane	ND	1.0	92	102	10.3				70 - 130	30
1,4-Dichlorobenzene	ND	1.0	91	98	7.4				70 - 130	30
1,4-dioxane	ND	100	96	94	2.1				70 - 130	30
2,2-Dichloropropane	ND	1.0	90	96	6.5				70 - 130	30
2-Chlorotoluene	ND	1.0	92	96	4.3				70 - 130	30
2-Hexanone	ND	5.0	71	80	11.9				70 - 130	30
2-Isopropyltoluene	ND	1.0	94	98	4.2				70 - 130	30
4-Chlorotoluene	ND	1.0	90	96	6.5				70 - 130	30
4-Methyl-2-pentanone	ND	5.0	76	90	16.9				70 - 130	30
Acetone	ND	5.0	63	73	14.7				70 - 130	30
Acrolein	ND	5.0	93	107	14.0				70 - 130	30
Acrylonitrile	ND	5.0	89	104	15.5				70 - 130	30
Benzene	ND	0.70	89	95	6.5				70 - 130	30
Bromobenzene	ND	1.0	92	98	6.3				70 - 130	30

QA/QC Data

SDG I.D.: GCC13976

Parameter	Blank		LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
	Blank	BLK RL								
Bromochloromethane	ND	1.0	88	100	12.8				70 - 130	30
Bromodichloromethane	ND	0.50	89	100	11.6				70 - 130	30
Bromoform	ND	1.0	94	107	12.9				70 - 130	30
Bromomethane	ND	1.0	78	84	7.4				70 - 130	30
Carbon Disulfide	ND	1.0	91	99	8.4				70 - 130	30
Carbon tetrachloride	ND	1.0	89	95	6.5				70 - 130	30
Chlorobenzene	ND	1.0	92	98	6.3				70 - 130	30
Chloroethane	ND	1.0	96	105	9.0				70 - 130	30
Chloroform	ND	1.0	86	93	7.8				70 - 130	30
Chloromethane	ND	1.0	76	85	11.2				70 - 130	30
cis-1,2-Dichloroethene	ND	1.0	90	97	7.5				70 - 130	30
cis-1,3-Dichloropropene	ND	0.40	92	101	9.3				70 - 130	30
Dibromochloromethane	ND	0.50	96	111	14.5				70 - 130	30
Dibromomethane	ND	1.0	89	102	13.6				70 - 130	30
Dichlorodifluoromethane	ND	1.0	98	107	8.8				70 - 130	30
Ethylbenzene	ND	1.0	93	98	5.2				70 - 130	30
Hexachlorobutadiene	ND	0.40	97	100	3.0				70 - 130	30
Isopropylbenzene	ND	1.0	91	95	4.3				70 - 130	30
m&p-Xylene	ND	1.0	91	97	6.4				70 - 130	30
Methyl ethyl ketone	ND	5.0	71	87	20.3				70 - 130	30
Methyl t-butyl ether (MTBE)	ND	1.0	91	107	16.2				70 - 130	30
Methylene chloride	ND	1.0	87	97	10.9				70 - 130	30
Naphthalene	ND	1.0	95	112	16.4				70 - 130	30
n-Butylbenzene	ND	1.0	94	97	3.1				70 - 130	30
n-Propylbenzene	ND	1.0	92	96	4.3				70 - 130	30
o-Xylene	ND	1.0	92	99	7.3				70 - 130	30
p-Isopropyltoluene	ND	1.0	94	96	2.1				70 - 130	30
sec-Butylbenzene	ND	1.0	96	99	3.1				70 - 130	30
Styrene	ND	1.0	92	99	7.3				70 - 130	30
tert-butyl alcohol	ND	10	104	110	5.6				70 - 130	30
tert-Butylbenzene	ND	1.0	91	95	4.3				70 - 130	30
Tetrachloroethene	ND	1.0	90	98	8.5				70 - 130	30
Tetrahydrofuran (THF)	ND	2.5	84	99	16.4				70 - 130	30
Toluene	ND	1.0	90	97	7.5				70 - 130	30
trans-1,2-Dichloroethene	ND	1.0	90	98	8.5				70 - 130	30
trans-1,3-Dichloropropene	ND	0.40	89	99	10.6				70 - 130	30
trans-1,4-dichloro-2-butene	ND	5.0	96	107	10.8				70 - 130	30
Trichloroethene	ND	1.0	92	97	5.3				70 - 130	30
Trichlorofluoromethane	ND	1.0	94	101	7.2				70 - 130	30
Trichlorotrifluoroethane	ND	1.0	93	101	8.2				70 - 130	30
Vinyl chloride	ND	1.0	94	101	7.2				70 - 130	30
% 1,2-dichlorobenzene-d4	100	%	99	101	2.0				70 - 130	30
% Bromofluorobenzene	95	%	100	102	2.0				70 - 130	30
% Dibromofluoromethane	101	%	100	103	3.0				70 - 130	30
% Toluene-d8	99	%	100	99	1.0				70 - 130	30

Comment:

A LCS and LCS Duplicate were performed instead of a matrix spike and matrix spike duplicate.

Additional 8260 criteria: 10% of LCS/LCSD compounds can be outside of acceptance criteria as long as recovery is 40-160%.

QA/QC Batch 460164 (ug/kg), QC Sample No: CC13978 (CC13976 (50X) , CC13978 (50X))

Volatiles - Soil

1,1,2,2-Tetrachloroethane	ND	3.0	116	111	4.4	107	106	0.9	70 - 130	30
1,2,3-Trichlorobenzene	ND	5.0	109	106	2.8	103	104	1.0	70 - 130	30

QA/QC Data

SDG I.D.: GCC13976

Parameter	Blank	Blk RL	LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
1,2,3-Trichloropropane	ND	5.0	117	115	1.7	111	110	0.9	70 - 130	30
1,2,4-Trichlorobenzene	ND	5.0	112	105	6.5	110	112	1.8	70 - 130	30
1,2,4-Trimethylbenzene	ND	1.0	108	108	0.0	109	110	0.9	70 - 130	30
1,2-Dibromo-3-chloropropane	ND	5.0	111	108	2.7	100	98	2.0	70 - 130	30
1,2-Dichlorobenzene	ND	5.0	111	110	0.9	109	109	0.0	70 - 130	30
1,3,5-Trimethylbenzene	ND	1.0	108	108	0.0	110	110	0.0	70 - 130	30
1,3-Dichlorobenzene	ND	5.0	112	109	2.7	111	111	0.0	70 - 130	30
1,4-Dichlorobenzene	ND	5.0	112	108	3.6	111	110	0.9	70 - 130	30
2-Chlorotoluene	ND	5.0	109	111	1.8	110	110	0.0	70 - 130	30
2-Isopropyltoluene	ND	5.0	113	112	0.9	114	114	0.0	70 - 130	30
4-Chlorotoluene	ND	5.0	109	109	0.0	110	109	0.9	70 - 130	30
Bromobenzene	ND	5.0	109	109	0.0	107	108	0.9	70 - 130	30
Hexachlorobutadiene	ND	5.0	111	107	3.7	117	118	0.9	70 - 130	30
Isopropylbenzene	ND	1.0	109	110	0.9	109	109	0.0	70 - 130	30
Naphthalene	ND	5.0	108	105	2.8	99	99	0.0	70 - 130	30
n-Butylbenzene	ND	1.0	113	108	4.5	116	116	0.0	70 - 130	30
n-Propylbenzene	ND	1.0	110	110	0.0	111	112	0.9	70 - 130	30
p-Isopropyltoluene	ND	1.0	111	109	1.8	114	114	0.0	70 - 130	30
sec-Butylbenzene	ND	1.0	116	114	1.7	117	117	0.0	70 - 130	30
tert-Butylbenzene	ND	1.0	109	109	0.0	109	110	0.9	70 - 130	30
trans-1,4-dichloro-2-butene	ND	5.0	118	117	0.9	111	111	0.0	70 - 130	30
% 1,2-dichlorobenzene-d4	100	%	99	99	0.0	99	100	1.0	70 - 130	30
% Bromofluorobenzene	97	%	99	99	0.0	99	100	1.0	70 - 130	30

Comment:

Additional 8260 criteria: 10% of LCS/LCSD compounds can be outside of acceptance criteria as long as recovery is 40-160%.

QA/QC Batch 460156 (ug/kg), QC Sample No: CC14920 (CC13976, CC13977, CC13978, CC13979, CC13980)

Volatiles - Soil

1,1,1,2-Tetrachloroethane	ND	5.0	102	102	0.0	112	110	1.8	70 - 130	30
1,1,1-Trichloroethane	ND	5.0	105	107	1.9	116	114	1.7	70 - 130	30
1,1,2,2-Tetrachloroethane	ND	3.0	105	104	1.0	109	109	0.0	70 - 130	30
1,1,2-Trichloroethane	ND	5.0	99	97	2.0	108	108	0.0	70 - 130	30
1,1-Dichloroethane	ND	5.0	104	104	0.0	113	112	0.9	70 - 130	30
1,1-Dichloroethene	ND	5.0	105	106	0.9	113	113	0.0	70 - 130	30
1,1-Dichloropropene	ND	5.0	107	108	0.9	119	118	0.8	70 - 130	30
1,2,3-Trichlorobenzene	ND	5.0	97	96	1.0	98	107	8.8	70 - 130	30
1,2,3-Trichloropropane	ND	5.0	103	103	0.0	112	112	0.0	70 - 130	30
1,2,4-Trichlorobenzene	ND	5.0	99	96	3.1	108	113	4.5	70 - 130	30
1,2,4-Trimethylbenzene	ND	1.0	102	102	0.0	111	111	0.0	70 - 130	30
1,2-Dibromo-3-chloropropane	ND	5.0	101	103	2.0	98	105	6.9	70 - 130	30
1,2-Dibromoethane	ND	5.0	100	101	1.0	110	109	0.9	70 - 130	30
1,2-Dichlorobenzene	ND	5.0	102	102	0.0	112	113	0.9	70 - 130	30
1,2-Dichloroethane	ND	5.0	105	104	1.0	117	116	0.9	70 - 130	30
1,2-Dichloropropane	ND	5.0	100	99	1.0	108	108	0.0	70 - 130	30
1,3,5-Trimethylbenzene	ND	1.0	102	102	0.0	111	111	0.0	70 - 130	30
1,3-Dichlorobenzene	ND	5.0	103	101	2.0	114	114	0.0	70 - 130	30
1,3-Dichloropropane	ND	5.0	102	102	0.0	111	108	2.7	70 - 130	30
1,4-Dichlorobenzene	ND	5.0	102	101	1.0	113	114	0.9	70 - 130	30
1,4-dioxane	ND	100	103	103	0.0	107	109	1.9	70 - 130	30
2,2-Dichloropropane	ND	5.0	110	110	0.0	122	121	0.8	70 - 130	30
2-Chlorotoluene	ND	5.0	104	104	0.0	111	112	0.9	70 - 130	30
2-Hexanone	ND	25	76	78	2.6	75	77	2.6	70 - 130	30
2-Isopropyltoluene	ND	5.0	106	106	0.0	107	107	0.0	70 - 130	30

QA/QC Data

SDG I.D.: GCC13976

Parameter	Blk		LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits	
	Blank	RL									
4-Chlorotoluene	ND	5.0	102	100	2.0	110	110	0.0	70 - 130	30	
4-Methyl-2-pentanone	ND	25	82	84	2.4	85	86	1.2	70 - 130	30	
Acetone	ND	10	61	62	1.6	45	45	0.0	70 - 130	30	l,m
Acrolein	ND	25	95	97	2.1	83	84	1.2	70 - 130	30	
Acrylonitrile	ND	5.0	97	100	3.0	99	98	1.0	70 - 130	30	
Benzene	ND	1.0	101	102	1.0	110	110	0.0	70 - 130	30	
Bromobenzene	ND	5.0	102	101	1.0	108	109	0.9	70 - 130	30	
Bromochloromethane	ND	5.0	105	105	0.0	115	114	0.9	70 - 130	30	
Bromodichloromethane	ND	5.0	101	102	1.0	112	111	0.9	70 - 130	30	
Bromoform	ND	5.0	98	97	1.0	105	103	1.9	70 - 130	30	
Bromomethane	ND	5.0	122	120	1.7	85	97	13.2	70 - 130	30	
Carbon Disulfide	ND	5.0	106	106	0.0	107	106	0.9	70 - 130	30	
Carbon tetrachloride	ND	5.0	92	107	15.1	101	113	11.2	70 - 130	30	
Chlorobenzene	ND	5.0	104	104	0.0	113	112	0.9	70 - 130	30	
Chloroethane	ND	5.0	116	114	1.7	124	123	0.8	70 - 130	30	
Chloroform	ND	5.0	104	103	1.0	113	111	1.8	70 - 130	30	
Chloromethane	ND	5.0	94	95	1.1	100	99	1.0	70 - 130	30	
cis-1,2-Dichloroethene	ND	5.0	106	105	0.9	113	109	3.6	70 - 130	30	
cis-1,3-Dichloropropene	ND	5.0	104	103	1.0	114	112	1.8	70 - 130	30	
Dibromochloromethane	ND	3.0	104	105	1.0	111	112	0.9	70 - 130	30	
Dibromomethane	ND	5.0	99	100	1.0	111	110	0.9	70 - 130	30	
Dichlorodifluoromethane	ND	5.0	121	123	1.6	134	132	1.5	70 - 130	30	m
Ethylbenzene	ND	1.0	104	104	0.0	114	113	0.9	70 - 130	30	
Hexachlorobutadiene	ND	5.0	102	104	1.9	117	117	0.0	70 - 130	30	
Isopropylbenzene	ND	1.0	104	105	1.0	111	110	0.9	70 - 130	30	
m&p-Xylene	ND	2.0	103	103	0.0	114	113	0.9	70 - 130	30	
Methyl ethyl ketone	ND	5.0	79	79	0.0	82	79	3.7	70 - 130	30	
Methyl t-butyl ether (MTBE)	ND	1.0	103	102	1.0	102	101	1.0	70 - 130	30	
Methylene chloride	ND	5.0	88	87	1.1	97	94	3.1	70 - 130	30	
Naphthalene	ND	5.0	97	97	0.0	91	102	11.4	70 - 130	30	
n-Butylbenzene	ND	1.0	104	105	1.0	116	117	0.9	70 - 130	30	
n-Propylbenzene	ND	1.0	105	106	0.9	112	112	0.0	70 - 130	30	
o-Xylene	ND	2.0	103	103	0.0	114	112	1.8	70 - 130	30	
p-Isopropyltoluene	ND	1.0	105	105	0.0	114	114	0.0	70 - 130	30	
sec-Butylbenzene	ND	1.0	109	110	0.9	118	117	0.9	70 - 130	30	
Styrene	ND	5.0	102	101	1.0	113	112	0.9	70 - 130	30	
tert-butyl alcohol	ND	100	105	101	3.9	91	101	10.4	70 - 130	30	
tert-Butylbenzene	ND	1.0	104	104	0.0	111	110	0.9	70 - 130	30	
Tetrachloroethene	ND	5.0	103	104	1.0	116	117	0.9	70 - 130	30	
Tetrahydrofuran (THF)	ND	5.0	98	97	1.0	96	95	1.0	70 - 130	30	
Toluene	ND	1.0	101	101	0.0	112	111	0.9	70 - 130	30	
trans-1,2-Dichloroethene	ND	5.0	105	106	0.9	116	112	3.5	70 - 130	30	
trans-1,3-Dichloropropene	ND	5.0	101	100	1.0	113	111	1.8	70 - 130	30	
trans-1,4-dichloro-2-butene	ND	5.0	109	108	0.9	103	104	1.0	70 - 130	30	
Trichloroethene	ND	5.0	106	108	1.9	117	117	0.0	70 - 130	30	
Trichlorofluoromethane	ND	5.0	115	115	0.0	114	112	1.8	70 - 130	30	
Trichlorotrifluoroethane	ND	5.0	105	107	1.9	107	105	1.9	70 - 130	30	
Vinyl chloride	ND	5.0	111	110	0.9	115	113	1.8	70 - 130	30	
% 1,2-dichlorobenzene-d4	100	%	101	99	2.0	100	101	1.0	70 - 130	30	
% Bromofluorobenzene	97	%	100	98	2.0	101	100	1.0	70 - 130	30	
% Dibromofluoromethane	97	%	96	95	1.0	97	97	0.0	70 - 130	30	
% Toluene-d8	99	%	100	99	1.0	100	100	0.0	70 - 130	30	

QA/QC Data

SDG I.D.: GCC13976

Parameter	Blank	Blk RL	LCS %	LCSD %	LCS RPD	MS %	MSD %	MS RPD	% Rec Limits	% RPD Limits
-----------	-------	-----------	----------	-----------	------------	---------	----------	-----------	--------------------	--------------------

Comment:

Additional 8260 criteria: 10% of LCS/LCSD compounds can be outside of acceptance criteria as long as recovery is 40-160%.

l = This parameter is outside laboratory LCS/LCSD specified recovery limits.

m = This parameter is outside laboratory MS/MSD specified recovery limits.

r = This parameter is outside laboratory RPD specified recovery limits.

If there are any questions regarding this data, please call Phoenix Client Services at extension 200.

RPD - Relative Percent Difference

LCS - Laboratory Control Sample

LCSD - Laboratory Control Sample Duplicate

MS - Matrix Spike

MS Dup - Matrix Spike Duplicate

NC - No Criteria

Intf - Interference



Phyllis Shiller, Laboratory Director

December 19, 2018

Wednesday, December 19, 2018

Criteria: NY: 375, 375GWP, 375RRS, 375RS, GW

State: NY

Sample Criteria Exceedances Report

GCC13976 - EBC

SampNo	Acode	Phoenix Analyte	Criteria	Result	RL	Criteria	Criteria	RL	Analysis Units
CC13976	\$8270SMRDP	Benz(a)anthracene	NY / 375-6.8 Semivolatiles / Ground Water Protection	3400	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Benzo(b)fluoranthene	NY / 375-6.8 Semivolatiles / Ground Water Protection	3000	260	1700	1700	1700	ug/Kg
CC13976	\$8270SMRDP	Chrysene	NY / 375-6.8 Semivolatiles / Ground Water Protection	3400	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Benzo(k)fluoranthene	NY / 375-6.8 Semivolatiles / Ground Water Protection	2700	260	1700	1700	1700	ug/Kg
CC13976	\$8270SMRDP	Dibenz(a,h)anthracene	NY / 375-6.8 Semivolatiles / Residential	720	190	330	330	330	ug/Kg
CC13976	\$8270SMRDP	Benzo(a)anthracene	NY / 375-6.8 Semivolatiles / Residential	3400	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Benzo(a)pyrene	NY / 375-6.8 Semivolatiles / Residential	3200	190	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Indeno(1,2,3-cd)pyrene	NY / 375-6.8 Semivolatiles / Residential	2100	260	500	500	500	ug/Kg
CC13976	\$8270SMRDP	Benzo(b)fluoranthene	NY / 375-6.8 Semivolatiles / Residential	3000	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Chrysene	NY / 375-6.8 Semivolatiles / Residential	3400	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Benzo(k)fluoranthene	NY / 375-6.8 Semivolatiles / Residential	2700	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Indeno(1,2,3-cd)pyrene	NY / 375-6.8 Semivolatiles / Residential Restricted	2100	260	500	500	500	ug/Kg
CC13976	\$8270SMRDP	Dibenz(a,h)anthracene	NY / 375-6.8 Semivolatiles / Residential Restricted	720	190	330	330	330	ug/Kg
CC13976	\$8270SMRDP	Benzo(b)fluoranthene	NY / 375-6.8 Semivolatiles / Residential Restricted	3000	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Benzo(a)anthracene	NY / 375-6.8 Semivolatiles / Residential Restricted	3400	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Benzo(a)pyrene	NY / 375-6.8 Semivolatiles / Residential Restricted	3200	190	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Indeno(1,2,3-cd)pyrene	NY / 375-6.8 Semivolatiles / Unrestricted Use Soil	2100	260	500	500	500	ug/Kg
CC13976	\$8270SMRDP	Benzo(b)fluoranthene	NY / 375-6.8 Semivolatiles / Unrestricted Use Soil	3000	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Benzo(a)pyrene	NY / 375-6.8 Semivolatiles / Unrestricted Use Soil	3200	190	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Chrysene	NY / 375-6.8 Semivolatiles / Unrestricted Use Soil	3400	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Dibenz(a,h)anthracene	NY / 375-6.8 Semivolatiles / Unrestricted Use Soil	720	190	330	330	330	ug/Kg
CC13976	\$8270SMRDP	Benzo(a)anthracene	NY / 375-6.8 Semivolatiles / Unrestricted Use Soil	3400	260	1000	1000	1000	ug/Kg
CC13976	\$8270SMRDP	Benzo(k)fluoranthene	NY / 375-6.8 Semivolatiles / Unrestricted Use Soil	2700	260	800	800	800	ug/Kg
CC13976	AS-SM	Arsenic	NY / 375-6.8 Metals / Unrestricted Use Soil	15.8	0.75	13	13	13	mg/Kg
CC13976	CD-SM	Cadmium	NY / 375-6.8 Metals / Residential	3.44	0.37	2.5	2.5	2.5	mg/Kg
CC13976	CD-SM	Cadmium	NY / 375-6.8 Metals / Unrestricted Use Soil	3.44	0.37	2.5	2.5	2.5	mg/Kg
CC13976	CR-SM	Chromium	NY / 375-6.8 Metals / Unrestricted Use Soil	46.3	0.37	30			mg/Kg
CC13976	CU-SM	Copper	NY / 375-6.8 Metals / Residential	308	7.5	270	270	270	mg/kg
CC13976	CU-SM	Copper	NY / 375-6.8 Metals / Residential Restricted	308	7.5	270	270	270	mg/kg
CC13976	CU-SM	Copper	NY / 375-6.8 Metals / Unrestricted Use Soil	308	7.5	50	50	50	mg/kg
CC13976	HG-SM	Mercury	NY / 375-6.8 Metals / Ground Water Protection	1.29	0.15	0.73	0.73	0.73	mg/Kg
CC13976	HG-SM	Mercury	NY / 375-6.8 Metals / Residential	1.29	0.15	0.81	0.81	0.81	mg/Kg
CC13976	HG-SM	Mercury	NY / 375-6.8 Metals / Residential Restricted	1.29	0.15	0.81	0.81	0.81	mg/Kg
CC13976	HG-SM	Mercury	NY / 375-6.8 Metals / Unrestricted Use Soil	1.29	0.15	0.18	0.18	0.18	mg/Kg
CC13976	NI-SM	Nickel	NY / 375-6.8 Metals / Unrestricted Use Soil	38.3	0.37	30	30	30	mg/Kg
CC13976	PB-SMDP	Lead	NY / 375-6.8 Metals / Ground Water Protection	662	7.5	450	450	450	mg/Kg
CC13976	PB-SMDP	Lead	NY / 375-6.8 Metals / Residential	662	7.5	400	400	400	mg/Kg
CC13976	PB-SMDP	Lead	NY / 375-6.8 Metals / Residential Restricted	662	7.5	400	400	400	mg/Kg
CC13976	PB-SMDP	Lead	NY / 375-6.8 Metals / Unrestricted Use Soil	662	7.5	63	63	63	mg/Kg
CC13976	ZN-SMDP	Zinc	NY / 375-6.8 Metals / Unrestricted Use Soil	720	7.5	109	109	109	mg/Kg
CC13977	CD-SM	Cadmium	NY / 375-6.8 Metals / Residential	2.61	0.40	2.5	2.5	2.5	mg/Kg

Wednesday, December 19, 2018

Criteria: NY: 375, 375GWP, 375RRS, 375RS, GW

State: NY

Sample Criteria Exceedances Report

GCC13976 - EBC

SampNo	Acode	Phoenix Analyte	Criteria	Result	RL	Criteria	RL Criteria	Analysis Units
CC13977	CD-SM	Cadmium	NY / 375-6.8 Metals / Unrestricted Use Soil	2.61	0.40	2.5	2.5	mg/Kg
CC13977	CU-SM	Copper	NY / 375-6.8 Metals / Unrestricted Use Soil	159	0.8	50	50	mg/kg
CC13977	HG-SM	Mercury	NY / 375-6.8 Metals / Ground Water Protection	13.0	1.4	0.73	0.73	mg/Kg
CC13977	HG-SM	Mercury	NY / 375-6.8 Metals / Residential	13.0	1.4	0.81	0.81	mg/Kg
CC13977	HG-SM	Mercury	NY / 375-6.8 Metals / Residential Restricted	13.0	1.4	0.81	0.81	mg/Kg
CC13977	HG-SM	Mercury	NY / 375-6.8 Metals / Unrestricted Use Soil	13.0	1.4	0.18	0.18	mg/Kg
CC13977	PB-SMDP	Lead	NY / 375-6.8 Metals / Ground Water Protection	959	8.0	450	450	mg/Kg
CC13977	PB-SMDP	Lead	NY / 375-6.8 Metals / Residential	959	8.0	400	400	mg/Kg
CC13977	PB-SMDP	Lead	NY / 375-6.8 Metals / Residential Restricted	959	8.0	400	400	mg/Kg
CC13977	PB-SMDP	Lead	NY / 375-6.8 Metals / Unrestricted Use Soil	959	8.0	63	63	mg/Kg
CC13977	ZN-SMDP	Zinc	NY / 375-6.8 Metals / Unrestricted Use Soil	877	8.0	109	109	mg/Kg
CC13978	CR-SM	Chromium	NY / 375-6.8 Metals / Unrestricted Use Soil	34.8	0.37	30		mg/Kg
CC13978	CU-SM	Copper	NY / 375-6.8 Metals / Unrestricted Use Soil	63.9	0.7	50	50	mg/kg
CC13978	HG-SM	Mercury	NY / 375-6.8 Metals / Unrestricted Use Soil	0.53	0.14	0.18	0.18	mg/Kg
CC13978	NI-SM	Nickel	NY / 375-6.8 Metals / Unrestricted Use Soil	43.2	0.37	30	30	mg/Kg
CC13978	PB-SMDP	Lead	NY / 375-6.8 Metals / Unrestricted Use Soil	382	7.4	63	63	mg/Kg
CC13978	ZN-SMDP	Zinc	NY / 375-6.8 Metals / Unrestricted Use Soil	195	7.4	109	109	mg/Kg
CC13979	HG-SM	Mercury	NY / 375-6.8 Metals / Unrestricted Use Soil	0.24	0.13	0.18	0.18	mg/Kg
CC13981	\$8260DP25R	1,2-Dibromoethane	NY / TOGS - Water Quality / GA Criteria	ND	0.25	0.0006	0.0006	ug/L
CC13981	\$8260DP25R	1,2,3-Trichloropropane	NY / TOGS - Water Quality / GA Criteria	ND	0.25	0.04	0.04	ug/L
CC13981	\$8260DP25R	1,2-Dibromo-3-chloropropane	NY / TOGS - Water Quality / GA Criteria	ND	0.50	0.04	0.04	ug/L
CC13982	\$8260DP25R	1,2,3-Trichloropropane	NY / TOGS - Water Quality / GA Criteria	ND	0.25	0.04	0.04	ug/L
CC13982	\$8260DP25R	1,2-Dibromo-3-chloropropane	NY / TOGS - Water Quality / GA Criteria	ND	0.50	0.04	0.04	ug/L
CC13982	\$8260DP25R	1,2-Dibromoethane	NY / TOGS - Water Quality / GA Criteria	ND	0.25	0.0006	0.0006	ug/L
CC13983	\$8260DP25R	1,2-Dibromoethane	NY / TOGS - Water Quality / GA Criteria	ND	0.25	0.0006	0.0006	ug/L
CC13983	\$8260DP25R	1,2,3-Trichloropropane	NY / TOGS - Water Quality / GA Criteria	ND	0.25	0.04	0.04	ug/L
CC13983	\$8260DP25R	1,2-Dibromo-3-chloropropane	NY / TOGS - Water Quality / GA Criteria	ND	0.50	0.04	0.04	ug/L

Phoenix Laboratories does not assume responsibility for the data contained in this exceedance report. It is provided as an additional tool to identify requested criteria exceedances. All efforts are made to ensure the accuracy of the data (obtained from appropriate agencies). A lack of exceedance information does not necessarily suggest conformance to the criteria. It is ultimately the site professional's responsibility to determine appropriate compliance.



Environmental Laboratories, Inc.
587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
Tel. (860) 645-1102 Fax (860) 645-0823



Analysis Comments

December 19, 2018

SDG I.D.: GCC13976

The following analysis comments are made regarding exceptions to criteria not already noted in the Analysis Report or QA/QC Report:

SVOA Narration

CHEM19 12/13/18-3: CC13976, CC13977, CC13978, CC13979, CC13980

The following Initial Calibration compounds did not meet RSD% criteria: 2,4-Dinitrophenol 26% (20%), 4,6-Dinitro-2-methylphenol 22% (20%)

The following Initial Calibration compounds did not meet maximum RSD% criteria: None.

The following Initial Calibration compounds did not meet recommended response factors: 2-Nitrophenol 0.054 (0.1), Hexachlorobenzene 0.090 (0.1)

The following Initial Calibration compounds did not meet minimum response factors: None.

The following Continuing Calibration compounds did not meet recommended response factors: 2-Nitrophenol 0.056 (0.1), Hexachlorobenzene 0.086 (0.1)

The following Continuing Calibration compounds did not meet minimum response factors: None.

Up to eight compounds can be outside of ICAL %RSD criteria and up to sixteen compounds can be outside of CCAL %Dev criteria if less than 40%.

VOA Narration

CHEM02 12/14/18-2: CC13982

The following Initial Calibration compounds did not meet RSD% criteria: Acetone 29% (20%)

The following Initial Calibration compounds did not meet maximum RSD% criteria: None.

The following Initial Calibration compounds did not meet recommended response factors: 1,2-Dibromo-3-chloropropane 0.028 (0.05), 2-Hexanone 0.085 (0.1), Acetone 0.057 (0.1), Acrolein 0.021 (0.05), Bromoform 0.084 (0.1), Methyl ethyl ketone 0.070 (0.1), Tetrahydrofuran (THF) 0.042 (0.05)

The following Initial Calibration compounds did not meet minimum response factors: None.

The following Continuing Calibration compounds did not meet % deviation criteria: Bromomethane 35%L (30%)

The following Continuing Calibration compounds did not meet Maximum % deviation criteria: None.

The following Continuing Calibration compounds did not meet recommended response factors: 1,1,2,2-Tetrachloroethane 0.242 (0.3), 1,2-Dibromo-3-chloropropane 0.029 (0.05), Acrolein 0.019 (0.05), Acrylonitrile 0.049 (0.05), Bromoform 0.083 (0.1), Tetrahydrofuran (THF) 0.038 (0.05)

The following Continuing Calibration compounds did not meet minimum response factors: None.

Up to eight compounds can be outside of ICAL %RSD criteria and up to sixteen compounds can be outside of CCAL %Dev criteria if less than 40%.

CHEM02 12/15/18-1: CC13981, CC13983

The following Initial Calibration compounds did not meet RSD% criteria: Acetone 29% (20%)

The following Initial Calibration compounds did not meet maximum RSD% criteria: None.

The following Initial Calibration compounds did not meet recommended response factors: 1,2-Dibromo-3-chloropropane 0.028 (0.05), 2-Hexanone 0.085 (0.1), Acetone 0.057 (0.1), Acrolein 0.021 (0.05), Bromoform 0.084 (0.1), Methyl ethyl ketone 0.070 (0.1), Tetrahydrofuran (THF) 0.042 (0.05)

The following Initial Calibration compounds did not meet minimum response factors: None.

The following Continuing Calibration compounds did not meet % deviation criteria: Bromomethane 33%L (30%)

The following Continuing Calibration compounds did not meet Maximum % deviation criteria: None.

The following Continuing Calibration compounds did not meet recommended response factors: 1,1,2,2-Tetrachloroethane 0.242 (0.3), 1,2-Dibromo-3-chloropropane 0.030 (0.05), Acrolein 0.021 (0.05), Bromoform 0.089 (0.1), Tetrahydrofuran (THF) 0.037 (0.05)

The following Continuing Calibration compounds did not meet minimum response factors: None.

Up to eight compounds can be outside of ICAL %RSD criteria and up to sixteen compounds can be outside of CCAL %Dev criteria if less than 40%.



Environmental Laboratories, Inc.
587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
Tel. (860) 645-1102 Fax (860) 645-0823



Analysis Comments

December 19, 2018

SDG I.D.: GCC13976

CHEM03 12/14/18-2: CC13976, CC13977, CC13978, CC13979, CC13980

The following Initial Calibration compounds did not meet RSD% criteria: 1,2,3-Trichlorobenzene 25% (20%), Acetone 36% (20%), Methyl Ethyl Ketone 21% (20%), Naphthalene 33% (20%)

The following Initial Calibration compounds did not meet maximum RSD% criteria: None.

Up to eight compounds can be outside of ICAL %RSD criteria and up to sixteen compounds can be outside of CCAL %Dev criteria if less than 40%.

CHEM03 12/15/18-1: CC13976, CC13978

The following Initial Calibration compounds did not meet RSD% criteria: 1,2,3-Trichlorobenzene 25% (20%), Naphthalene 33% (20%)

The following Initial Calibration compounds did not meet maximum RSD% criteria: None.

Up to eight compounds can be outside of ICAL %RSD criteria and up to sixteen compounds can be outside of CCAL %Dev criteria if less than 40%.



Environmental Laboratories, Inc.
587 East Middle Turnpike, P.O.Box 370, Manchester, CT 06045
Tel. (860) 645-1102 Fax (860) 645-0823



NY Temperature Narration

December 19, 2018

SDG I.D.: GCC13976

The samples in this delivery group were received at 3.6°C.
(Note acceptance criteria for relevant matrices is above freezing up to 6°C)

